

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122121\
 Data File : VD071414.D
 Acq On : 21 Dec 2021 16:53
 Operator : VA/SY
 Sample : M5184-06
 Misc : 7.68G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-6

Quant Time: Dec 21 19:09:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	94521	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	174623	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	170641	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	78938	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	90582	81.672	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	163.340%#
35) Dibromofluoromethane	7.914	113	66356	59.674	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	119.340%
50) Toluene-d8	10.338	98	218498	51.634	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.260%
62) 4-Bromofluorobenzene	12.626	95	77338	52.724	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122121\
Data File : VD071414.D
Acq On : 21 Dec 2021 16:53
Operator : VA/SY
Sample : M5184-06
Misc : 7.68G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-6

Quant Time: Dec 21 19:09:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

