

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122121\
 Data File : VD071419.D
 Acq On : 21 Dec 2021 19:12
 Operator : VA/SY
 Sample : M5182-04
 Misc : 5.94G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HB-3-(10-15)-12-17-21

Quant Time: Dec 21 20:01:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	459786	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	825292	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	699167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	191524	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	278124	51.552	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.100%
35) Dibromofluoromethane	7.914	113	259215	49.324	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.640%
50) Toluene-d8	10.332	98	1022082	51.106	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.220%
62) 4-Bromofluorobenzene	12.626	95	287898	41.529	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.060%
Target Compounds						
						Qvalue
16) Acetone	4.162	43	32024	50.396	ug/l	94
17) Carbon Disulfide	4.409	76	54218	4.338	ug/l #	94
18) Methyl Acetate	4.726	43	9357	2.351	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122121\
Data File : VD071419.D
Acq On : 21 Dec 2021 19:12
Operator : VA/SY
Sample : M5182-04
Misc : 5.94G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HB-3-(10-15)-12-17-21

Quant Time: Dec 21 20:01:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

