

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122121\
 Data File : VD071421.D
 Acq On : 21 Dec 2021 20:08
 Operator : VA/SY
 Sample : M5182-05
 Misc : 6.45G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HB-2-(20-25)-12-17-21

Quant Time: Dec 21 21:12:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	484997	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	851215	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	760367	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	346104	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	273510	48.061	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	96.120%		
35) Dibromofluoromethane	7.914	113	266275	49.124	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	98.240%		
50) Toluene-d8	10.338	98	1069577	51.852	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	103.700%		
62) 4-Bromofluorobenzene	12.626	95	369087	51.619	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	103.240%		
Target Compounds						
16) Acetone	4.156	43	20700	42.807	ug/l	95
17) Carbon Disulfide	4.420	76	45986	3.488	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122121\
Data File : VD071421.D
Acq On : 21 Dec 2021 20:08
Operator : VA/SY
Sample : M5182-05
Misc : 6.45G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HB-2-(20-25)-12-17-21

Quant Time: Dec 21 21:12:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

