

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122121\
 Data File : VD071403.D
 Acq On : 21 Dec 2021 10:19
 Operator : VA/SY
 Sample : VD1221SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1221SBL01

Quant Time: Dec 21 19:03:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	278254	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	488222	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	443599	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	203188	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	179947	55.114	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.220%
35) Dibromofluoromethane	7.914	113	164814	53.013	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.020%
50) Toluene-d8	10.338	98	564810	47.739	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.480%
62) 4-Bromofluorobenzene	12.626	95	208982	50.957	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.920%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122121\
Data File : VD071403.D
Acq On : 21 Dec 2021 10:19
Operator : VA/SY
Sample : VD1221SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1221SBL01

Quant Time: Dec 21 19:03:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. -0.006 min
Lab File: VD071403.D
Acq: 21 Dec 2021 10:19




Instrument :
MSVOA_D
ClientSampleId :
VD1221SBL01


Tgt Ion:168 Resp: 278254
Ion Ratio Lower Upper
168 100
99 59.1 46.6 69.8



#33
1,2-Dichloroethane-d4
Concen: 55.114 ug/l
RT: 8.332 min Scan# 1090
Delta R.T. 0.006 min
Lab File: VD071403.D
Acq: 21 Dec 2021 10:19

Tgt Ion: 65 Resp: 179947
Ion Ratio Lower Upper
65 100
67 50.6 0.0 112.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.861 min Scan# 1180

Delta R.T. 0.000 min

Lab File: VD071403.D

Acq: 21 Dec 2021 10:19

Instrument :

MSVOA_D

ClientSampleId :

VD1221SBL01





Tgt Ion:114 Resp: 488222

Ion Ratio Lower Upper

114 100

63 22.3 0.0 41.8

88 16.3 0.0 32.8



#35

Dibromofluoromethane

Concen: 53.013 ug/l

RT: 7.914 min Scan# 1019

Delta R.T. -0.006 min

Lab File: VD071403.D

Acq: 21 Dec 2021 10:19




Tgt Ion:113 Resp: 164814


Ion Ratio Lower Upper

113 100

111 104.1 82.0 123.0

192 18.0 15.2 22.8









#50
Toluene-d8
Concen: 47.739 ug/l
RT: 10.338 min Scan# 1431
Delta R.T. -0.000 min
Lab File: VD071403.D
Acq: 21 Dec 2021 10:19




Instrument :
MSVOA_D
ClientSampleId :
VD1221SBL01


Tgt Ion: 98 Resp: 564810
Ion Ratio Lower Upper
98 100
100 63.3 51.8 77.6



#62
4-Bromofluorobenzene
Concen: 50.957 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. -0.000 min
Lab File: VD071403.D
Acq: 21 Dec 2021 10:19

Tgt Ion: 95 Resp: 208982
Ion Ratio Lower Upper
95 100
174 76.6 0.0 159.4
176 71.7 0.0 154.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 117
Delta R.T. -0.000 min
Lab File: VD071403.D
Acq: 21 Dec 2021 10:19

Instrument :
MSVOA_D
ClientSampleId :
VD1221SBL01

Tgt Ion:117 Resp: 443599
Ion Ratio Lower Upper
117 100
82 56.7 44.7 67.1
119 31.8 26.2 39.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. -0.006 min
Lab File: VD071403.D
Acq: 21 Dec 2021 10:19

Tgt Ion:152 Resp: 203188
Ion Ratio Lower Upper
152 100
115 62.2 29.8 89.3
150 161.2 0.0 349.2

