

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122123\
 Data File : VD078022.D
 Acq On : 21 Dec 2023 13:17
 Operator : RP/MD
 Sample : VD1221SBL01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1221SBL01

Quant Time: Dec 22 05:11:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	188491	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	335909	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	316975	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	142888	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	83426	37.561	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	75.120%
35) Dibromofluoromethane	7.810	113	102167	43.421	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.840%
50) Toluene-d8	10.269	98	421201	44.695	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.400%
62) 4-Bromofluorobenzene	12.575	95	119969	42.580	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122123\
Data File : VD078022.D
Acq On : 21 Dec 2023 13:17
Operator : RP/MD
Sample : VD1221SBL01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1221SBL01

Quant Time: Dec 22 05:11:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

