

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122123\
 Data File : VD078030.D
 Acq On : 21 Dec 2023 17:12
 Operator : RP/MD
 Sample : 05877-01
 Misc : 5.01G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 COM07

Quant Time: Dec 22 05:24:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	169831	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	301378	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	274661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	110535	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	71968	35.962	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	71.920%
35) Dibromofluoromethane	7.810	113	89415	42.356	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.720%
50) Toluene-d8	10.269	98	381968	45.176	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.360%
62) 4-Bromofluorobenzene	12.575	95	98457	38.949	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122123\
Data File : VD078030.D
Acq On : 21 Dec 2023 17:12
Operator : RP/MD
Sample : 05877-01
Misc : 5.01G/5.0ml/MSVOA_D/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
COM07

Quant Time: Dec 22 05:24:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

