



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122223\
Data File : VD078055.D
Acq On : 22 Dec 2023 09:21
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Title : SW846 8260
Last Update : Thu Dec 14 14:48:10 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.4	10853	PASS
75	95	30	60	54.8	29112	PASS
95	95	100	100	100.0	53139	PASS
96	95	5	9	6.3	3338	PASS
173	174	0.00	2	1.8	887	PASS
174	95	50	100	90.8	48232	PASS
175	174	5	9	7.9	3809	PASS
176	174	95	101	95.7	46181	PASS
177	176	5	9	6.3	2923	PASS