

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122321\
 Data File : VD071448.D
 Acq On : 23 Dec 2021 17:41
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Dec 24 03:32:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	444047	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	782267	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	702632	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	300756	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	270870	45.931	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.860%
35) Dibromofluoromethane	7.914	113	247824	47.138	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.280%
50) Toluene-d8	10.332	98	982055	49.429	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.860%
62) 4-Bromofluorobenzene	12.620	95	316583	44.841	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.680%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122321\
Data File : VD071448.D
Acq On : 23 Dec 2021 17:41
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Dec 24 03:32:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. 0.006 min
Lab File: VD071448.D
Acq: 23 Dec 2021 17:41




Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Tgt Ion:168 Resp: 444047
Ion Ratio Lower Upper
168 100
99 60.7 46.6 69.8



#33
1,2-Dichloroethane-d4
Concen: 45.931 ug/l
RT: 8.326 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VD071448.D
Acq: 23 Dec 2021 17:41

Tgt Ion: 65 Resp: 270870
Ion Ratio Lower Upper
65 100
67 51.6 0.0 112.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.861 min Scan# 1180

Delta R.T. 0.000 min

Lab File: VD071448.D

Acq: 23 Dec 2021 17:41

Instrument :

MSVOA_D

ClientSampleId :

VIBLK




Tgt Ion:114 Resp: 782267

Ion Ratio Lower Upper

114 100

63 21.5 0.0 41.8

88 15.9 0.0 32.8



#35

Dibromofluoromethane

Concen: 47.138 ug/l

RT: 7.914 min Scan# 1019

Delta R.T. 0.006 min

Lab File: VD071448.D

Acq: 23 Dec 2021 17:41




Tgt Ion:113 Resp: 247824


Ion Ratio Lower Upper

113 100

111 103.1 82.0 123.0

192 17.4 15.2 22.8









#50
Toluene-d8
Concen: 49.429 ug/l
RT: 10.332 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VD071448.D
Acq: 23 Dec 2021 17:41




Instrument :
MSVOA_D
ClientSampleId :
VIBLK


Tgt Ion: 98 Resp: 982055
Ion Ratio Lower Upper
98 100
100 65.4 51.8 77.6



#62
4-Bromofluorobenzene
Concen: 44.841 ug/l
RT: 12.620 min Scan# 1819
Delta R.T. -0.006 min
Lab File: VD071448.D
Acq: 23 Dec 2021 17:41

Tgt Ion: 95 Resp: 316583
Ion Ratio Lower Upper
95 100
174 75.7 0.0 159.4
176 74.3 0.0 154.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 117
Delta R.T. -0.000 min
Lab File: VD071448.D
Acq: 23 Dec 2021 17:41

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Tgt Ion:117 Resp: 702632
Ion Ratio Lower Upper
117 100
82 56.7 44.7 67.1
119 30.7 26.2 39.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VD071448.D
Acq: 23 Dec 2021 17:41

Tgt Ion:152 Resp: 300756
Ion Ratio Lower Upper
152 100
115 61.5 29.8 89.3
150 158.9 0.0 349.2

