

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122521\
 Data File : VD071462.D
 Acq On : 25 Dec 2021 17:12
 Operator : VA/SY
 Sample : M5187-03
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SASP2-119-1-10-40

Quant Time: Dec 27 00:50:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	431887	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	754177	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	675069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	300752	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	253488	44.194	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.380%
35) Dibromofluoromethane	7.914	113	238191	46.993	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.980%
50) Toluene-d8	10.332	98	951075	49.653	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.300%
62) 4-Bromofluorobenzene	12.620	95	305588	44.896	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.800%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122521\
Data File : VD071462.D
Acq On : 25 Dec 2021 17:12
Operator : VA/SY
Sample : M5187-03
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SASP2-119-1-10-40

Quant Time: Dec 27 00:50:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

