

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122521\
 Data File : VD071464.D
 Acq On : 25 Dec 2021 18:08
 Operator : VA/SY
 Sample : M5191-01
 Misc : 6.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-(0-3.5)-12-21-2021

Quant Time: Dec 27 00:51:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	421713	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	736270	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	652209	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	279243	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	234733	41.911	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	83.820%
35) Dibromofluoromethane	7.908	113	221187	44.700	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.400%
50) Toluene-d8	10.332	98	914950	48.928	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.860%
62) 4-Bromofluorobenzene	12.626	95	295546	44.477	ug/l	0.00
Spiked Amount	50.000					

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122521\
Data File : VD071464.D
Acq On : 25 Dec 2021 18:08
Operator : VA/SY
Sample : M5191-01
Misc : 6.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-(0-3.5)-12-21-2021

Quant Time: Dec 27 00:51:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

