

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122521\
 Data File : VD071465.D
 Acq On : 25 Dec 2021 18:36
 Operator : VA/SY
 Sample : M5191-03
 Misc : 5.24G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-(3.5-7.0)-12-21-2021

Quant Time: Dec 27 00:51:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	17243	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	30421	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	29610	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	12634	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	16097	70.292	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	140.580%
35) Dibromofluoromethane	7.909	113	8168	39.951	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	79.900%
50) Toluene-d8	10.338	98	36213	46.870	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.740%
62) 4-Bromofluorobenzene	12.620	95	13064	47.583	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.160%
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	4262	74.095	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122521\
Data File : VD071465.D
Acq On : 25 Dec 2021 18:36
Operator : VA/SY
Sample : M5191-03
Misc : 5.24G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-(3.5-7.0)-12-21-2021

Quant Time: Dec 27 00:51:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

