

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122521\
 Data File : VD071471.D
 Acq On : 25 Dec 2021 21:24
 Operator : VA/SY
 Sample : M5223-03
 Misc : 5.56G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DB-4-(33-35)-12-22-21

Quant Time: Dec 27 00:52:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	351012	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	642467	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	589768	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	265570	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	220011	47.195	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.380%
35) Dibromofluoromethane	7.914	113	201075	46.568	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.140%
50) Toluene-d8	10.332	98	810925	49.697	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.400%
62) 4-Bromofluorobenzene	12.620	95	271133	46.760	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.520%
Target Compounds						
17) Carbon Disulfide	4.420	76	18382	1.543	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122521\
Data File : VD071471.D
Acq On : 25 Dec 2021 21:24
Operator : VA/SY
Sample : M5223-03
Misc : 5.56G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DB-4-(33-35)-12-22-21

Quant Time: Dec 27 00:52:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

