

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD122619\
 Data File : VD064595.D
 Acq On : 26 Dec 2019 13:14
 Operator : VA/SY
 Sample : K6435-02
 Misc : 6.10G/5.00ml/MSVOA D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-4E-(21.5-23.5)

Quant Time: Dec 27 02:52:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	467735	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	741035	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	783021	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	411073	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	203997	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
35) Dibromofluoromethane	7.92	113	225519	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
50) Toluene-d8	10.34	98	916235	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
62) 4-Bromofluorobenzene	12.64	95	356687	64.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	129.56%	
Target Compounds						
30) Chloroform	7.71	83	58197	7.220	ug/l	98
95) Naphthalene	15.41	128	18073	1.322	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD122619\
Data File : VD064595.D
Acq On : 26 Dec 2019 13:14
Operator : VA/SY
Sample : K6435-02
Misc : 6.10G/5.00ml/MSVOA D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-4E-(21.5-23.5)

Quant Time: Dec 27 02:52:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

