

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD122619\
 Data File : VD064597.D
 Acq On : 26 Dec 2019 14:12
 Operator : VA/SY
 Sample : K6436-01
 Misc : 5.32G/5.00ml/MSVOA D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 K009-SB-1

Quant Time: Dec 27 03:11:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	487445	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	788380	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	808996	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	406902	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	207059	48.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.34%	
35) Dibromofluoromethane	7.92	113	213029	44.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.08%	
50) Toluene-d8	10.34	98	961130	51.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.76%	
62) 4-Bromofluorobenzene	12.64	95	364275	62.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	124.38%	
Target Compounds						
16) Acetone	4.17	43	6452	10.466	ug/l #	72
25) 2-Butanone	7.23	43	22205	18.794	ug/l	97
30) Chloroform	7.71	83	17041	2.029	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD122619\
Data File : VD064597.D
Acq On : 26 Dec 2019 14:12
Operator : VA/SY
Sample : K6436-01
Misc : 5.32G/5.00ml/MSVOA D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
K009-SB-1

Quant Time: Dec 27 03:11:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

