

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD122619\
 Data File : VD064611.D
 Acq On : 26 Dec 2019 20:53
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 27 04:56:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.389	0.385	1.0	94	0.00
3 P	Chloromethane	0.543	0.576	-6.1	102	0.00
4 C	Vinyl Chloride	0.663	0.675	-1.8#	99	0.00
5 T	Bromomethane	0.482	0.503	-4.4	104	0.00
6 T	Chloroethane	0.444	0.475	-7.0	100	0.00
7 T	Trichlorofluoromethane	1.116	1.136	-1.8	97	0.00
8 T	Diethyl Ether	0.222	0.244	-9.9	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.462	0.462	0.0	94	0.00
10 T	Methyl Iodide	0.496	0.556	-12.1	97	0.00
11 T	Tert butyl alcohol	0.025	0.029	-16.0	101	0.00
12 CM	1,1-Dichloroethene	0.440	0.454	-3.2#	98	0.00
13 T	Acrolein	0.035	0.037	-5.7	96	0.00
14 T	Allyl chloride	0.686	0.745	-8.6	103	0.00
15 T	Acrylonitrile	0.092	0.107	-16.3	107	0.00
16 T	Acetone	0.091	0.083	8.8	87	0.00
17 T	Carbon Disulfide	1.422	1.461	-2.7	99	0.00
18 T	Methyl Acetate	0.248	0.271	-9.3	105	0.00
19 T	Methyl tert-butyl Ether	0.977	1.131	-15.8	107	0.00
20 T	Methylene Chloride	0.524	0.547	-4.4	109	0.00
21 T	trans-1,2-Dichloroethene	0.501	0.541	-8.0	104	0.00
22 T	Diisopropyl ether	1.368	1.585	-15.9	108	0.00
23 T	Vinyl Acetate	0.776	0.926	-19.3	107	0.00
24 P	1,1-Dichloroethane	0.821	0.922	-12.3	106	0.00
25 T	2-Butanone	0.121	0.133	-9.9	101	0.00
26 T	2,2-Dichloropropane	0.764	0.747	2.2	93	0.00
27 T	cis-1,2-Dichloroethene	0.541	0.609	-12.6	106	0.00
28 T	Bromochloromethane	0.319	0.346	-8.5	95	0.00
29 T	Tetrahydrofuran	0.075	0.086	-14.7	104	0.00
30 C	Chloroform	0.862	0.969	-12.4#	107	0.00
31 T	Cyclohexane	0.822	0.786	4.4	95	0.00
32 T	1,1,1-Trichloroethane	0.809	0.851	-5.2	100	0.00
33 S	1,2-Dichloroethane-d4	0.441	0.495	-12.2	105	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.307	0.327	-6.5	102	0.00
36 T	1,1-Dichloropropene	0.481	0.477	0.8	98	0.00
37 T	Ethyl Acetate	0.189	0.206	-9.0	103	0.00
38 T	Carbon Tetrachloride	0.513	0.521	-1.6	98	0.00
39 T	Methylcyclohexane	0.583	0.572	1.9	95	0.00
40 TM	Benzene	1.314	1.394	-6.1	105	0.00
41 T	Methacrylonitrile	0.110	0.118	-7.3	97	-0.01
42 TM	1,2-Dichloroethane	0.380	0.413	-8.7	107	0.00
43 T	Isopropyl Acetate	0.376	0.409	-8.8	104	0.00
44 TM	Trichloroethene	0.386	0.397	-2.8	102	0.00
45 C	1,2-Dichloropropane	0.314	0.344	-9.6#	108	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD122619\
 Data File : VD064611.D
 Acq On : 26 Dec 2019 20:53
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 27 04:56:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.175	0.192	-9.7	106	0.00
47 T	Bromodichloromethane	0.465	0.516	-11.0	108	0.00
48 T	Methyl methacrylate	0.177	0.189	-6.8	98	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	112	0.00
50 S	Toluene-d8	1.175	1.243	-5.8	102	0.00
51 T	4-Methyl-2-Pentanone	0.186	0.206	-10.8	107	0.00
52 CM	Toluene	0.862	0.920	-6.7#	105	0.00
53 T	t-1,3-Dichloropropene	0.434	0.475	-9.4	105	0.00
54 T	cis-1,3-Dichloropropene	0.512	0.556	-8.6	104	0.00
55 T	1,1,2-Trichloroethane	0.236	0.254	-7.6	106	0.00
56 T	Ethyl methacrylate	0.294	0.334	-13.6	108	0.00
57 T	1,3-Dichloropropane	0.406	0.451	-11.1	108	0.00
58 T	2-Chloroethyl Vinyl ether	0.112	0.110	1.8	101	0.00
59 T	2-Hexanone	0.126	0.136	-7.9	101	0.00
60 T	Dibromochloromethane	0.326	0.363	-11.3	108	0.00
61 T	1,2-Dibromoethane	0.232	0.257	-10.8	107	0.00
62 S	4-Bromofluorobenzene	0.372	0.401	-7.8	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
64 T	Tetrachloroethene	0.384	0.373	2.9	101	0.00
65 PM	Chlorobenzene	1.038	1.080	-4.0	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.391	0.421	-7.7	110	0.00
67 C	Ethyl Benzene	1.884	1.929	-2.4#	104	0.00
68 T	m/p-Xylenes	0.727	0.744	-2.3	103	0.00
69 T	o-Xylene	0.657	0.689	-4.9	106	0.00
70 T	Styrene	1.153	1.237	-7.3	108	0.00
71 P	Bromoform	0.223	0.243	-9.0	110	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.573	3.594	-0.6	101	0.00
74 T	N-amyl acetate	0.764	0.826	-8.1	107	0.00
75 P	1,1,2,2-Tetrachloroethane	0.561	0.595	-6.1	108	0.00
76 T	1,2,3-Trichloropropane	0.395	0.427	-8.1	129	0.00
77 T	Bromobenzene	0.878	0.923	-5.1	107	0.00
78 T	n-propylbenzene	4.168	4.200	-0.8	102	0.00
79 T	2-Chlorotoluene	2.290	2.362	-3.1	106	0.00
80 T	1,3,5-Trimethylbenzene	2.935	3.005	-2.4	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.189	0.195	-3.2	105	0.00
82 T	4-Chlorotoluene	2.437	2.515	-3.2	106	0.00
83 T	tert-Butylbenzene	2.548	2.628	-3.1	104	0.00
84 T	1,2,4-Trimethylbenzene	2.923	3.037	-3.9	105	0.00
85 T	sec-Butylbenzene	3.494	3.505	-0.3	101	0.00
86 T	p-Isopropyltoluene	3.341	3.366	-0.7	101	0.00
87 T	1,3-Dichlorobenzene	1.674	1.747	-4.4	107	0.00
88 T	1,4-Dichlorobenzene	1.644	1.703	-3.6	105	0.00
89 T	n-Butylbenzene	3.008	3.001	0.2	100	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD122619\
Data File : VD064611.D
Acq On : 26 Dec 2019 20:53
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Dec 27 04:56:41 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.638	0.642	-0.6	102	0.00
91 T	1,2-Dichlorobenzene	1.427	1.495	-4.8	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.091	0.093	-2.2	107	0.00
93 T	1,2,4-Trichlorobenzene	1.051	1.100	-4.7	106	0.00
94 T	Hexachlorobutadiene	0.661	0.646	2.3	100	0.00
95 T	Naphthalene	1.662	1.831	-10.2	108	0.00
96 T	1,2,3-Trichlorobenzene	0.897	0.963	-7.4	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6