

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD122619\
 Data File : VD064616.D
 Acq On : 27 Dec 2019 00:37
 Operator : VA/SY
 Sample : VD1226SBS02
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1226SBS02

Manual Integrations
 APPROVED

apatel
 12/27/2019 2:07:21 PM

Quant Time: Dec 27 05:57:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	371306	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	571346	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	517716	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	247613	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.33	65	192387	58.76	ug/l	0.00
Spiked Amount			Recovery	=	117.52%	
35) Dibromofluoromethane	7.92	113	196572	56.08	ug/l	0.00
Spiked Amount			Recovery	=	112.16%	
50) Toluene-d8	10.34	98	738434	55.00	ug/l	0.00
Spiked Amount			Recovery	=	110.00%	
62) 4-Bromofluorobenzene	12.64	95	233343	54.97	ug/l	0.00
Spiked Amount			Recovery	=	109.94%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.99	85	57496	19.925	ug/l	97
3) Chloromethane	2.21	50	86578	21.453	ug/l	97
4) Vinyl Chloride	2.35	62	99689	20.260	ug/l	97
5) Bromomethane	2.76	94	79924	22.339	ug/l	98
6) Chloroethane	2.92	64	71946	21.821	ug/l	100
7) Trichlorofluoromethane	3.27	101	168088	20.288	ug/l	94
8) Diethyl Ether	3.70	74	37748	22.944	ug/l	98
9) 1,1,2-Trichlorotrifluoroet	4.09	101	69290	20.194	ug/l	99
10) Methyl Iodide	4.29	142	78612	21.329	ug/l	99
11) Tert butyl alcohol	5.25	59	21113	112.518	ug/l #	91
12) 1,1-Dichloroethene	4.06	96	67072	20.532	ug/l	97
13) Acrolein	3.92	56	27117	103.695	ug/l	97
14) Allyl chloride	4.70	41	109514	21.502	ug/l	100
15) Acrylonitrile	5.43	53	77748	113.451	ug/l	99
16) Acetone	4.17	43	64448	89.389	ug/l	89
17) Carbon Disulfide	4.40	76	216667	20.512	ug/l	100
18) Methyl Acetate	4.73	43	43941	23.838	ug/l	96
19) Methyl tert-butyl Ether	5.49	73	161640	22.277	ug/l	96
20) Methylene Chloride	4.96	84	86531	23.224	ug/l	97
21) trans-1,2-Dichloroethene	5.47	96	80571	21.657	ug/l	96
22) Diisopropyl ether	6.37	45	227787	22.423	ug/l	99
23) Vinyl Acetate	6.31	43	608919	105.666	ug/l	98
24) 1,1-Dichloroethane	6.27	63	136702	22.415	ug/l	98
25) 2-Butanone	7.22	43	95594	106.218	ug/l	97
26) 2,2-Dichloropropane	7.21	77	109669	19.341	ug/l	99
27) cis-1,2-Dichloroethene	7.21	96	91157	22.704	ug/l	97
28) Bromochloromethane	7.55	49	47716	20.127	ug/l	98
29) Tetrahydrofuran	7.57	42	61642	110.139	ug/l	98
30) Chloroform	7.71	83	144911	22.648	ug/l	95
31) Cyclohexane	7.99	56	117509	19.249	ug/l #	93
32) 1,1,1-Trichloroethane	7.91	97	126373	21.027	ug/l	98
36) 1,1-Dichloropropene	8.11	75	108972	19.816	ug/l	100
37) Ethyl Acetate	7.30	43	45816	21.208	ug/l	99
38) Carbon Tetrachloride	8.10	117	114608	19.545	ug/l	90

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD122619\
 Data File : VD064616.D
 Acq On : 27 Dec 2019 00:37
 Operator : VA/SY
 Sample : VD1226SBS02
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleID :
 VD1226SBS02

Manual Integrations
 APPROVED

apatel
 12/27/2019 2:07:21 PM

Quant Time: Dec 27 05:57:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	9.36	83	123753	18.580	ug/l	97
40) Benzene	8.35	78	307303	20.470	ug/l	100
41) Methacrylonitrile	7.53	41	25109	20.032	ug/l	91
42) 1,2-Dichloroethane	8.43	62	89818	20.704	ug/l	99
43) Isopropyl Acetate	8.46	43	88261	20.526	ug/l	100
44) Trichloroethene	9.11	130	89489	20.292	ug/l	98
45) 1,2-Dichloropropane	9.39	63	76326	21.294	ug/l	99
46) Dibromomethane	9.47	93	43095	21.565	ug/l	98
47) Bromodichloromethane	9.66	83	113570	21.357	ug/l #	99
48) Methyl methacrylate	9.46	41	39791	19.653	ug/l	87
49) 1,4-Dioxane	9.46	88	10090	437.879	ug/l	93
51) 4-Methyl-2-Pentanone	10.23	43	216780	101.994	ug/l	99
52) Toluene	10.40	92	197439	20.046	ug/l	99
53) t-1,3-Dichloropropene	10.62	75	99086	19.986	ug/l	95
54) cis-1,3-Dichloropropene	10.09	75	119586	20.456	ug/l	98
55) 1,1,2-Trichloroethane	10.80	97	56990	21.161	ug/l	95
56) Ethyl methacrylate	10.66	69	70035	20.822	ug/l	99
57) 1,3-Dichloropropane	10.95	76	97820	21.095	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.94	63	141410	110.352	ug/l	100
59) 2-Hexanone	10.98	43	142647	98.885	ug/l	99
60) Dibromochloromethane	11.14	129	76625	20.574	ug/l	99
61) 1,2-Dibromoethane	11.25	107	55875	21.035	ug/l	97
64) Tetrachloroethene	10.87	164	76234	19.194	ug/l	96
65) Chlorobenzene	11.67	112	217198	20.214	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.74	131	80745	19.933	ug/l	99
67) Ethyl Benzene	11.75	91	376788	19.319	ug/l	99
68) m/p-Xylenes	11.86	106	286054	38.000	ug/l	98
69) o-Xylene	12.18	106	133291	19.603	ug/l	99
70) Styrene	12.20	104	236699	19.832	ug/l	98
71) Bromoform	12.36	173	46246	20.020	ug/l #	99
73) Isopropylbenzene	12.48	105	351321	19.854	ug/l	99
74) N-amyl acetate	12.29	43	75349	19.920	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.74	83	58596	21.080	ug/l	98
76) 1,2,3-Trichloropropane	12.79	75	41747m	21.351	ug/l	
77) Bromobenzene	12.77	156	91658	21.081	ug/l	97
78) n-propylbenzene	12.83	91	411337	19.926	ug/l	100
79) 2-Chlorotoluene	12.91	91	230664	20.340	ug/l	99
80) 1,3,5-Trimethylbenzene	12.96	105	290150	19.961	ug/l	99
81) trans-1,4-Dichloro-2-buten	12.53	75	18387	19.667	ug/l	99
82) 4-Chlorotoluene	13.01	91	245140	20.313	ug/l	100
83) tert-Butylbenzene	13.23	119	255350	20.233	ug/l	99
84) 1,2,4-Trimethylbenzene	13.27	105	294874	20.369	ug/l	99
85) sec-Butylbenzene	13.41	105	335416	19.386	ug/l	99
86) p-Isopropyltoluene	13.52	119	319547	19.315	ug/l	99
87) 1,3-Dichlorobenzene	13.52	146	165861	20.011	ug/l	99
88) 1,4-Dichlorobenzene	13.60	146	169451	20.811	ug/l	99
89) n-Butylbenzene	13.85	91	284026	19.064	ug/l	100
90) Hexachloroethane	14.12	117	63706	20.159	ug/l	97
91) 1,2-Dichlorobenzene	13.90	146	147030	20.803	ug/l	99
92) 1,2-Dibromo-3-Chloropropan	14.51	75	9649	21.413	ug/l	91

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD122619\
Data File : VD064616.D
Acq On : 27 Dec 2019 00:37
Operator : VA/SY
Sample : VD1226SBSD02
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBSD02


Manual Integrations
APPROVED

apatel
12/27/2019 2:07:21 PM

Quant Time: Dec 27 05:57:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	15.17	180	102679	19.731	ug/l	99
94) Hexachlorobutadiene	15.28	225	64397	19.665	ug/l	97
95) Naphthalene	15.41	128	165326	20.081	ug/l	99
96) 1,2,3-Trichlorobenzene	15.60	180	90323	20.331	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.98 min Scan# 1030

Delta R.T. -0.01 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBS02

Tot Ion:168 Resp: 371306

Ion Ratio Lower Upper

168 100


99 49.7 41.8 62.6

Manual Integrations

APPROVED


apatel

12/27/2019 2:07:21 PM





Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC



Time-->



#2

Dichlorodifluoromethane

Concen: 19.925 ug/l

RT: 1.99 min Scan# 12

Delta R.T. 0.00 min

Lab File: VD064616.D


Acq: 27 Dec 2019 00:37

Tgt Ion: 85 Resp: 57496

Ion Ratio Lower Upper


85 100

87 30.3 16.1 48.2





Abundance Ion 84.90 (84.60 to 85.60): VDC

Ion 86.90 (86.60 to 87.60): VDC



Time-->





#3

Chloromethane

Concen: 21.453 ug/l

RT: 2.21 min Scan# 49

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBS02

Tot Ion: 50 Resp: 86578

Ion Ratio Lower Upper

50 100


52 34.8 26.5 39.7

Manual Integrations

APPROVED




apatel

12/27/2019 2:07:21 PM



Abundance Ion 49.90 (49.60 to 50.60): VDC

Ion 51.90 (51.60 to 52.60): VDC



#4

Vinyl Chloride

Concen: 20.260 ug/l

RT: 2.35 min Scan# 73

Delta R.T. 0.00 min

Lab File: VD064616.D


Acq: 27 Dec 2019 00:37

Tgt Ion: 62 Resp: 99689

Ion Ratio Lower Upper



62 100


64 31.4 26.6 40.0



Abundance Ion 61.90 (61.60 to 62.60): VDC

Ion 63.90 (63.60 to 64.60): VDC





#5

Bromomethane

Concen: 22.339 ug/l

RT: 2.76 min Scan# 143

Delta R.T. -0.01 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBS02

Tot Ion: 94 Resp: 79924

Ion Ratio Lower Upper

94 100


96 94.4 77.2 115.8

Manual Integrations

APPROVED

apatel

12/27/2019 2:07:21 PM



Abundance Ion 93.90 (93.60 to 94.60): VDC

Ion 95.90 (95.60 to 96.60): VDC

2.76



30000

20000

10000

0

Time-->



#6

Chloroethane

Concen: 21.821 ug/l

RT: 2.92 min Scan# 170

Delta R.T. 0.00 min

Lab File: VD064616.D


Acq: 27 Dec 2019 00:37

Tgt Ion: 64 Resp: 71946

Ion Ratio Lower Upper

64 100

66 32.6 25.9 38.9



Abundance Ion 63.90 (63.60 to 64.60): VDC

Ion 65.90 (65.60 to 66.60): VDC

2.92


30000


20000

10000

0

Time-->





#7





Trichlorofluoromethane
Concen: 20.288 ug/l
RT: 3.27 min Scan# 229
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBSD02

Tot Ion	Ratio	Lower	Upper
101	100		
103	60.2	52.2	78.4

Manual Integrations
APPROVED




apatel
12/27/2019 2:07:21 PM




#8

Diethyl Ether
Concen: 22.944 ug/l
RT: 3.70 min Scan# 303
Delta R.T. -0.01 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Lower	Upper
74	100		
45	93.3	47.5	142.5





#9

1,1,2-Trichlorotrifluoroethane

Concen: 20.194 ug/l

RT: 4.09 min Scan# 368

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D





ClientSampleId :

VD1226SBSD02

Tot Ion:	101	Resp:	69290
Ion Ratio	Lower	Upper	
101	100		
85	44.4	35.2	52.8
151	83.9	66.2	99.4

Manual Integrations
APPROVED

apatel
12/27/2019 2:07:21 PM



#10

Methyl Iodide

Concen: 21.329 ug/l




RT: 4.29 min Scan# 403

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Tgt Ion:	142	Resp:	78612
Ion Ratio	Lower	Upper	
142	100		
127	46.1	36.2	54.4
141	14.0	11.8	17.8





#11





Tert butyl alcohol
Concen: 112.518 ug/l
RT: 5.25 min Scan# 566
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBSD02

Tot Ion	59	Resp	21113
Ion Ratio	100	Lower	Upper
57	8.0	9.3	13.9#

Manual Integrations
APPROVED




apatel
12/27/2019 2:07:21 PM




#12

1,1-Dichloroethene
Concen: 20.532 ug/l
RT: 4.06 min Scan# 363
Delta R.T. -0.01 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Tgt Ion	96	Resp	67072
Ion Ratio	100	Lower	Upper
61	164.5	129.4	194.0
98	61.3	51.4	77.0





#13

Acrolein

Concen: 103.695 ug/l

RT: 3.92 min Scan# 340

Delta R.T. -0.01 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBS02

Tot Ion: 56 Resp: 27117

Ion Ratio Lower Upper

56 100





55 68.6 56.7 85.1

Manual Integrations

APPROVED

apatel

12/27/2019 2:07:21 PM



#14

Allyl chloride

Concen: 21.502 ug/l

RT: 4.70 min Scan# 473

Delta R.T. -0.01 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37




Tgt Ion: 41 Resp: 109514


Ion Ratio Lower Upper

41 100

39 78.0 62.3 93.5

76 37.8 29.7 44.5





#15

Acrylonitrile

Concen: 113.451 ug/l

RT: 5.43 min Scan# 597

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D





ClientSampleId :

VD1226SBS02

Tot Ion:	53	Resp:	77748
Ion Ratio	Lower	Upper	
53	100		
52	80.9	66.2	99.2
51	37.6	30.1	45.1

Manual Integrations
APPROVED

apatel
12/27/2019 2:07:21 PM



#16

Acetone

Concen: 89.389 ug/l




RT: 4.17 min Scan# 382


Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Tgt Ion:	43	Resp:	64448
Ion Ratio	Lower	Upper	
43	100		
58	25.8	25.8	38.6









#17
Carbon Disulfide
Concen: 20.512 ug/l
RT: 4.40 min Scan# 421
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBSD02

Tot Ion	Ratio	Lower	Upper
76	100		
78	9.1	7.3	10.9




Manual Integrations
APPROVED


apatel
12/27/2019 2:07:21 PM



#18
Methyl Acetate
Concen: 23.838 ug/l
RT: 4.73 min Scan# 477
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Lower	Upper
43	100		
74	25.8	19.2	28.8









#19
Methyl tert-butyl Ether
Concen: 22.277 ug/l
RT: 5.49 min Scan# 606
Delta R.T. 0.01 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBS02

Tot Ion: 73 Resp: 161640
Ion Ratio Lower Upper
73 100
57 25.3 18.6 27.8




Manual Integrations
APPROVED


apatel
12/27/2019 2:07:21 PM



#20
Methylene Chloride
Concen: 23.224 ug/l
RT: 4.96 min Scan# 517
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Tgt Ion: 84 Resp: 86531
Ion Ratio Lower Upper
84 100
49 124.6 97.2 145.8
51 34.8 31.4 47.0
86 63.5 49.3 73.9





#21

trans-1,2-Dichloroethene

Concen: 21.657 ug/l

RT: 5.47 min Scan# 603

Delta R.T. -0.01 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D


ClientSampleId :

VD1226SBS02

Tot Ion:	96	Resp:	80571
Ion	Ratio	Lower	Upper
96	100		
61	135.4	104.7	157.1
98	66.4	49.4	74.2

Manual Integrations
APPROVED

apatel
12/27/2019 2:07:21 PM






Abundance

Ion 95.90 (95.60 to 96.60): VDC

Ion 60.90 (60.60 to 61.60): VDC

Ion 97.90 (97.60 to 98.60): VDC



#22

Diisopropyl ether

Concen: 22.423 ug/l

RT: 6.37 min Scan# 757

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Tgt Ion:	45	Resp:	227787
Ion	Ratio	Lower	Upper
45	100		
43	53.9	43.2	64.8
87	26.7	21.9	32.9
59	10.9	9.0	13.4




Abundance


Ion 45.00 (44.70 to 45.70): VDC

Ion 43.00 (42.70 to 43.70): VDC

Ion 87.00 (86.70 to 87.70): VDC

Ion 59.00 (58.70 to 59.70): VDC





#23

Vinyl Acetate

Concen: 105.666 ug/l

RT: 6.31 min Scan# 746

Delta R.T. -0.01 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBS02

Tot Ion: 43 Resp: 608919

Ion Ratio Lower Upper

43 100


86 11.1 8.3 12.5

Manual Integrations

APPROVED

apatel

12/27/2019 2:07:21 PM



Abundance Ion 43.00 (42.70 to 43.70): VDC

Ion 86.00 (85.70 to 86.70): VDC

6.31

200000



150000

100000

50000

0

Time-->



#24

1,1-Dichloroethane

Concen: 22.415 ug/l

RT: 6.27 min Scan# 739

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37


Tgt Ion: 63 Resp: 136702

Ion Ratio Lower Upper

63 100

98 9.0 4.1 12.3

100 4.1 2.2 6.6



Abundance Ion 62.90 (62.60 to 63.60): VDC

Ion 97.90 (97.60 to 98.60): VDC

Ion 99.90 (99.60 to 100.60): VDC

6.27

60000

50000

40000


30000


20000

10000

0

Time-->





#25

2-Butanone

Concen: 106.218 ug/l

RT: 7.22 min Scan# 901

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBS02

Tot Ion: 43 Resp: 95594

Ion Ratio Lower Upper

43 100





72 24.5 20.7 31.1

Manual Integrations

APPROVED

apatel

12/27/2019 2:07:21 PM



#26

2,2-Dichloropropane

Concen: 19.341 ug/l

RT: 7.21 min Scan# 899

Delta R.T. 0.00 min

Lab File: VD064616.D




Acq: 27 Dec 2019 00:37


Tgt Ion: 77 Resp: 109669

Ion Ratio Lower Upper

77 100

97 22.7 11.2 33.6





#27





cis-1,2-Dichloroethene
 Concen: 22.704 ug/l
 RT: 7.21 min Scan# 900
 Delta R.T. 0.00 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1226SBS02

Tot Ion	Ratio	Lower	Upper
96	100		
61	137.9	0.0	282.6
98	63.5	0.0	130.8

Manual Integrations
 APPROVED




apatel
 12/27/2019 2:07:21 PM




#28

Bromochloromethane
 Concen: 20.127 ug/l
 RT: 7.55 min Scan# 957
 Delta R.T. 0.00 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Lower	Upper
49	100		
129	3.0	0.0	4.8
130	82.7	65.0	97.6









#29
Tetrahydrofuran
Concen: 110.139 ug/l
RT: 7.57 min Scan# 960
Delta R.T. -0.01 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBS02

Tot Ion	Ratio	Lower	Upper
42	100		
72	43.0	35.0	52.4
71	39.4	32.9	49.3




Manual Integrations
APPROVED


apatel
12/27/2019 2:07:21 PM



#30
Chloroform
Concen: 22.648 ug/l
RT: 7.71 min Scan# 985
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Lower	Upper
83	100		
85	68.1	51.1	76.7









#31
Cyclohexane
Concen: 19.249 ug/l
RT: 7.99 min Scan# 1031
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBS02

Tot Ion: 56 Resp: 117509
Ion Ratio Lower Upper
56 100
69 38.3 24.7 37.1#
84 91.1 69.4 104.2




Manual Integrations
APPROVED


apatel
12/27/2019 2:07:21 PM



#32
1,1,1-Trichloroethane
Concen: 21.027 ug/l
RT: 7.91 min Scan# 1018
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Tgt Ion: 97 Resp: 126373
Ion Ratio Lower Upper
97 100
99 62.4 51.0 76.6
61 45.0 35.4 53.0





#33

1,2-Dichloroethane-d4

Concen: 58.760 ug/l

RT: 8.33 min Scan# 1090

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBSD02

Tot Ion: 65 Resp: 192387

Ion Ratio Lower Upper

65 100





67 52.2 0.0 106.2

Manual Integrations

APPROVED

apatel

12/27/2019 2:07:21 PM



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.87 min Scan# 1181

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37




Tgt Ion: 114 Resp: 571346

Ion Ratio Lower Upper

114 100

63 18.4 0.0 38.0

88 14.7 0.0 29.2









#35
 Dibromofluoromethane
 Concen: 56.078 ug/l
 RT: 7.92 min Scan# 1020
 Delta R.T. 0.00 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1226SBS02

Tot Ion	Ratio	Resp	Lower	Upper
113	100	196572		
111	102.5	80.6	121.0	
192	21.8	17.9	26.9	




Manual Integrations
 APPROVED


apatel
 12/27/2019 2:07:21 PM



#36
 1,1-Dichloropropene
 Concen: 19.816 ug/l
 RT: 8.11 min Scan# 1052
 Delta R.T. -0.01 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Resp	Lower	Upper
75	100	108972		
110	36.1	18.1	54.1	
77	30.7	24.2	36.4	









#37
Ethyl Acetate
Concen: 21.208 ug/l
RT: 7.30 min Scan# 915
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBS02

Tot Ion	Ratio	Lower	Upper
43	100		
61	12.9	10.3	15.5
70	10.3	7.6	11.4




Manual Integrations
APPROVED

apatel
12/27/2019 2:07:21 PM



#38
Carbon Tetrachloride
Concen: 19.545 ug/l
RT: 8.10 min Scan# 1051
Delta R.T. 0.01 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Lower	Upper
117	100		
119	86.7	77.3	115.9
121	26.3	25.0	37.6









#39
Methvlcvclohexane
Concen: 18.580 ug/l
RT: 9.36 min Scan# 1264
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBSD02

Tot Ion	Ratio	Lower	Upper
83	100		
55	73.9	57.8	86.6
98	51.7	38.7	58.1




Manual Integrations
APPROVED


apatel
12/27/2019 2:07:21 PM



#40
Benzene
Concen: 20.470 ug/l
RT: 8.35 min Scan# 1093
Delta R.T. -0.01 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Lower	Upper
78	100		
77	24.9	19.9	29.9






#41
Methacrylonitrile
Concen: 20.032 ug/l
RT: 7.53 min Scan# 953
Delta R.T. -0.01 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBS02




Tot Ion:	41	Resp:	25109
Ion	Ratio	Lower	Upper
41	100		
39	56.9	46.7	70.1
67	77.0	52.5	78.7
52	34.9	24.1	36.1

Manual Integrations
APPROVED

apatel
12/27/2019 2:07:21 PM




Abundance Ion 41.00 (40.70 to 41.70): VDC
Ion 39.00 (38.70 to 39.70): VDC
Ion 67.00 (66.70 to 67.70): VDC
Ion 52.00 (51.70 to 52.70): VDC






#42
1,2-Dichloroethane
Concen: 20.704 ug/l
RT: 8.43 min Scan# 1106
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Tgt Ion:	62	Resp:	89818
Ion	Ratio	Lower	Upper
62	100		
98	10.4	0.0	20.0



Abundance Ion 61.90 (61.60 to 62.60): VDC
Ion 97.90 (97.60 to 98.60): VDC





#43

Isopropyl Acetate

Concen: 20.526 ug/l

RT: 8.46 min Scan# 1111

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D





ClientSampleId :

VD1226SBS02

Tot Ion:	43	Resp:	88261
Ion	Ratio	Lower	Upper
43	100		
61	29.6	23.7	35.5
87	13.1	10.5	15.7

Manual Integrations
APPROVED

apatel
12/27/2019 2:07:21 PM



#44

Trichloroethene

Concen: 20.292 ug/l




RT: 9.11 min Scan# 1223


Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Tgt Ion:	130	Resp:	89489
Ion	Ratio	Lower	Upper
130	100		
95	93.3	0.0	190.2





#45

1,2-Dichloropropane

Concen: 21.294 ug/l

RT: 9.39 min Scan# 1270

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBSD02

Tot Ion: 63 Resp: 76326

Ion Ratio Lower Upper

63 100





65 30.8 25.2 37.8

Manual Integrations

APPROVED

apatel

12/27/2019 2:07:21 PM



#46

Dibromomethane

Concen: 21.565 ug/l

RT: 9.47 min Scan# 1284

Delta R.T. -0.01 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37




Tgt Ion: 93 Resp: 43095


Ion Ratio Lower Upper

93 100

95 83.6 64.6 97.0

174 108.6 88.1 132.1





#47

Bromodichloromethane

Concen: 21.357 ug/l

RT: 9.66 min Scan# 1316

Delta R.T. -0.01 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D





ClientSampleId :

VD1226SBS02

Tot Ion:	83	Resp:	113570
Ion	Ratio	Lower	Upper
83	100		
85	62.4	49.5	74.3
127	7.5	7.8	11.6#

Manual Integrations
APPROVED

apatel
12/27/2019 2:07:21 PM



#48

Methyl methacrylate

Concen: 19.653 ug/l




RT: 9.46 min Scan# 1282


Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Tgt Ion:	41	Resp:	39791
Ion	Ratio	Lower	Upper
41	100		
69	90.6	63.0	94.6
39	68.9	47.4	71.2









#49
 1,4-Dioxane
 Concen: 437.879 ug/l
 RT: 9.46 min Scan# 1282
 Delta R.T. 0.00 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1226SBSD02

Tot Ion	88	Resp	10090
Ion Ratio	Lower	Upper	
88	100		
43	37.6	27.7	41.5
58	60.2	53.8	80.6




Manual Integrations
 APPROVED


apatel
 12/27/2019 2:07:21 PM



#50
 Toluene-d8
 Concen: 55.005 ug/l
 RT: 10.34 min Scan# 1431
 Delta R.T. 0.00 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Tgt Ion	98	Resp	738434
Ion Ratio	Lower	Upper	
98	100		
100	65.6	51.7	77.5





#51

4-Methyl-2-Pentanone

Concen: 101.994 ug/l

RT: 10.23 min Scan# 1413

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBS02

Tot Ion: 43 Resp: 216780

Ion Ratio Lower Upper

43 100


58 38.2 30.2 45.2

Manual Integrations

APPROVED




apatel

12/27/2019 2:07:21 PM



Abundance Ion 43.00 (42.70 to 43.70): VDC

Ion 58.00 (57.70 to 58.70): VDC



#52

Toluene

Concen: 20.046 ug/l

RT: 10.40 min Scan# 1442

Delta R.T. 0.00 min

Lab File: VD064616.D


Acq: 27 Dec 2019 00:37

Tgt Ion: 92 Resp: 197439

Ion Ratio Lower Upper



92 100


91 174.7 138.7 208.1



Abundance Ion 92.00 (91.70 to 92.70): VDC

Ion 91.00 (90.70 to 91.70): VDC









#53
 t-1,3-Dichloropropene
 Concen: 19.986 ug/l
 RT: 10.62 min Scan# 1479
 Delta R.T. 0.00 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1226SBS02

Tot Ion: 75 Resp: 99086
 Ion Ratio Lower Upper
 75 100
 77 29.1 25.4 38.2




Manual Integrations
 APPROVED


apatel
 12/27/2019 2:07:21 PM



#54
 cis-1,3-Dichloropropene
 Concen: 20.456 ug/l
 RT: 10.09 min Scan# 1389
 Delta R.T. 0.00 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Tgt Ion: 75 Resp: 119586
 Ion Ratio Lower Upper
 75 100
 77 31.1 24.6 36.8
 39 49.5 41.4 62.2





#55

1,1,2-Trichloroethane

Concen: 21.161 ug/l

RT: 10.80 min Scan# 1510

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D





ClientSampleId :

VD1226SBS02

Tot Ion:	97	Resp:	56990
Ion	Ratio	Lower	Upper
97	100		
83	80.3	70.2	105.4
85	53.6	44.0	66.0
99	61.5	47.8	71.6

Manual Integrations
APPROVED

apatel
12/27/2019 2:07:21 PM



#56

Ethyl methacrylate

Concen: 20.822 ug/l




RT: 10.66 min Scan# 1486


Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Tgt Ion:	69	Resp:	70035
Ion	Ratio	Lower	Upper
69	100		
41	68.6	55.3	82.9
39	41.3	33.8	50.8





#57

1,3-Dichloropropane

Concen: 21.095 ug/l

RT: 10.95 min Scan# 1535

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBSD02

Tot Ion: 76 Resp: 97820

Ion Ratio Lower Upper

76 100


78 32.0 26.2 39.4

Manual Integrations

APPROVED

apatel

12/27/2019 2:07:21 PM



Abundance Ion 75.90 (75.60 to 76.60): VDC

Ion 77.90 (77.60 to 78.60): VDC

10.95

60000

50000

40000



30000

20000

10000

0

Time-->



#58

2-Chloroethyl Vinyl ether

Concen: 110.352 ug/l

RT: 9.94 min Scan# 1364

Delta R.T. 0.00 min

Lab File: VD064616.D


Acq: 27 Dec 2019 00:37

Tgt Ion: 63 Resp: 141410

Ion Ratio Lower Upper

63 100

106 30.0 23.8 35.8



Abundance Ion 62.90 (62.60 to 63.60): VDC

Ion 105.90 (105.60 to 106.60): V

9.94

80000


60000


40000

20000

0

Time-->









#59
2-Hexanone
Concen: 98.885 ug/l
RT: 10.98 min Scan# 1541
Delta R.T. -0.01 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBS02

Tot Ion: 43 Resp: 142647
Ion Ratio Lower Upper
43 100
58 53.4 26.4 79.0




Manual Integrations
APPROVED


apatel
12/27/2019 2:07:21 PM



#60
Dibromochloromethane
Concen: 20.574 ug/l
RT: 11.14 min Scan# 1568
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Tgt Ion: 129 Resp: 76625
Ion Ratio Lower Upper
129 100
127 78.0 38.4 115.2









#61
 1,2-Dibromoethane
 Concen: 21.035 ug/l
 RT: 11.25 min Scan# 1586
 Delta R.T. 0.00 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1226SBSD02

Tot Ion	Ratio	Lower	Upper
107	100		
109	91.5	75.6	113.4




Manual Integrations
 APPROVED


apatel
 12/27/2019 2:07:21 PM



#62
 4-Bromofluorobenzene
 Concen: 54.967 ug/l
 RT: 12.64 min Scan# 1822
 Delta R.T. 0.00 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Lower	Upper
95	100		
174	96.9	0.0	193.2
176	91.4	0.0	184.2






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.65 min Scan# 1654
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBS02

Tot Ion	Ratio	Lower	Upper
117	100		
82	51.2	42.2	63.4
119	31.3	26.0	39.0

Manual Integrations
APPROVED

apatel
12/27/2019 2:07:21 PM






Abundance

Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): V

Ion 118.90 (118.60 to 119.60): V



#64
Tetrachloroethene
Concen: 19.194 ug/l
RT: 10.87 min Scan# 1522
Delta R.T. -0.01 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Lower	Upper
164	100		
166	123.1	101.3	151.9
129	82.2	70.5	105.7
131	79.1	66.6	99.8




Abundance

Ion 163.80 (163.50 to 164.50): V

Ion 165.80 (165.50 to 166.50): V

Ion 128.80 (128.50 to 129.50): V

Ion 130.80 (130.50 to 131.50): V





#65

Chlorobenzene

Concen: 20.214 ug/l

RT: 11.67 min Scan# 1658

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBSD02

Tot Ion:112 Resp: 217198

Ion Ratio Lower Upper

112 100





114 33.2 24.8 37.2

Manual Integrations

APPROVED

apatel

12/27/2019 2:07:21 PM



#66

1,1,1,2-Tetrachloroethane

Concen: 19.933 ug/l

RT: 11.74 min Scan# 1670

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37




Tgt Ion:131 Resp: 80745

Ion Ratio Lower Upper

131 100

133 95.7 48.3 144.8

119 66.0 32.1 96.3





#67

Ethyl Benzene

Concen: 19.319 ug/l

RT: 11.75 min Scan# 1671

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBS02

Tot Ion: 91 Resp: 376788

Ion Ratio Lower Upper

91 100


106 31.1 24.6 36.8

Manual Integrations




APPROVED

apatel

12/27/2019 2:07:21 PM



Abundance Ion 91.00 (90.70 to 91.70): VDC



#68

m/p-Xylenes

Concen: 38.000 ug/l

RT: 11.86 min Scan# 1689

Delta R.T. 0.00 min

Lab File: VD064616.D


Acq: 27 Dec 2019 00:37

Tgt Ion: 106 Resp: 286054



Ion Ratio Lower Upper


106 100

91 198.1 156.3 234.5



Abundance Ion 106.00 (105.70 to 106.70): VDC









#69
o-Xylene
Concen: 19.603 ug/l
RT: 12.18 min Scan# 1745
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBS02

Tot Ion:106 Resp: 133291
Ion Ratio Lower Upper
106 100
91 208.1 104.8 314.4




Manual Integrations
APPROVED


apatel
12/27/2019 2:07:21 PM



#70
Styrene
Concen: 19.832 ug/l
RT: 12.20 min Scan# 1747
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Tgt Ion:104 Resp: 236699
Ion Ratio Lower Upper
104 100
78 47.5 38.2 57.2
103 53.8 44.6 67.0





#71

Bromoform

Concen: 20.020 ug/l

RT: 12.36 min Scan# 1775

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBS02

Tot Ion:173 Resp: 46246

Ion Ratio Lower Upper

173 100

175 48.8 24.6 73.8


254 0.0 0.1 0.1#

Manual Integrations

APPROVED

apatel




12/27/2019 2:07:21 PM



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.58 min Scan# 1982

Delta R.T. -0.01 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37


Tgt Ion:152 Resp: 247613

Ion Ratio Lower Upper

152 100

115 53.0 27.8 83.4



150 164.0 0.0 351.4




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 19.854 ug/l

RT: 12.48 min Scan# 1796

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBS02

Tot Ion: 105 Resp: 351321

Ion Ratio Lower Upper

105 100


120 27.6 13.5 40.5

Manual Integrations

APPROVED




apatel

12/27/2019 2:07:21 PM



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#74

N-ethyl acetate

Concen: 19.920 ug/l

RT: 12.29 min Scan# 1763

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Tgt Ion: 43 Resp: 75349


Ion Ratio Lower Upper

43 100

70 42.3 33.4 50.0

55 26.3 19.4 29.0

61 22.9 18.3 27.5






Abundance Ion 43.00 (42.70 to 43.70): VDC

Ion 70.00 (69.70 to 70.70): VDC

Ion 55.00 (54.70 to 55.70): VDC

Ion 61.00 (60.70 to 61.70): VDC





#75

1,1,2,2-Tetrachloroethane

Concen: 21.080 ug/l

RT: 12.74 min Scan# 1839

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D


ClientSampleId :

VD1226SBS02




Tot Ion:	83	Resp:	58596
Ion Ratio	Lower	Upper	
83	100		
131	11.7	5.8	17.4
85	63.1	32.3	96.9

Manual Integrations
APPROVED

apatel
12/27/2019 2:07:21 PM



Abundance Ion 82.90 (82.60 to 83.60): VDC
Ion 130.80 (130.50 to 131.50): VDC
Ion 84.90 (84.60 to 85.60): VDC



#76

1,2,3-Trichloropropane

Concen: 21.351 ug/l m


RT: 12.79 min Scan# 1848

Delta R.T. 0.00 min



Lab File: VD064616.D


Acq: 27 Dec 2019 00:37

Tgt Ion:	75	Resp:	41747
Ion Ratio	Lower	Upper	
75	100		
77	0.0	0.0	0.0



Abundance Ion 74.90 (74.60 to 75.60): VDC
Ion 77.00 (76.70 to 77.70): VDC





#77

Bromobenzene

Concen: 21.081 ug/l

RT: 12.77 min Scan# 1844

Delta R.T. 0.01 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D


ClientSampleId :

VD1226SBSD02

Tot Ion	Ion	Ratio	Resp	Lower	Upper
156	100				
77	156.0	80.9	242.7		
158	95.8	48.0	144.2		

Manual Integrations
APPROVED




apatel
12/27/2019 2:07:21 PM



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): V

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 19.926 ug/l


RT: 12.83 min Scan# 1854

Delta R.T. 0.00 min

Lab File: VD064616.D



Acq: 27 Dec 2019 00:37

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
91	100				
120	24.0	11.9	35.9		



Abundance Ion 91.00 (90.70 to 91.70): V

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 20.340 ug/l

RT: 12.91 min Scan# 1868

Delta R.T. -0.01 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D

ClientSampleId :

VD1226SBSD02

Tot Ion: 91 Resp: 230664

Ion Ratio Lower Upper

91 100

126 36.6

17.8


53.5

Manual Integrations




APPROVED

apatel

12/27/2019 2:07:21 PM



Abundance Ion 91.00 (90.70 to 91.70): VD064584.D (-1862) (-)



#80

1,3,5-Trimethylbenzene

Concen: 19.961 ug/l

RT: 12.96 min Scan# 1877

Delta R.T. -0.01 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Tgt Ion:105 Resp: 290150


Ion Ratio Lower Upper

105 100



120 50.7


25.0

75.0



Abundance Ion 105.00 (104.70 to 105.70): VD064584.D (-1868) (-)





#81

trans-1,4-Dichloro-2-butene

Concen: 19.667 ug/l

RT: 12.53 min Scan# 1804

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D





ClientSampleId :

VD1226SBSD02

Tot Ion:	75	Resp:	18387
Ion	Ratio	Lower	Upper
75	100		
53	98.3	78.8	118.2
89	46.7	38.3	57.5

Manual Integrations
APPROVED

apatel
12/27/2019 2:07:21 PM



#82

4-Chlorotoluene

Concen: 20.313 ug/l




RT: 13.01 min Scan# 1885


Delta R.T. -0.01 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Tgt Ion:	91	Resp:	245140
Ion	Ratio	Lower	Upper
91	100		
126	36.0	17.8	53.5





#83

tert-Butylbenzene

Concen: 20.233 ug/l

RT: 13.23 min Scan# 1922

Delta R.T. -0.01 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37

Instrument :

MSVOA_D


ClientSampleId :

VD1226SBS02

Tot Ion:	119	Resp:	255350
Ion Ratio	Lower	Upper	
119	100		
91	61.2	30.8	92.4
134	27.4	12.8	38.3

Manual Integrations
APPROVED

apatel
12/27/2019 2:07:21 PM






Abundance

Ion 119.00 (118.70 to 119.70): V

Ion 91.00 (90.70 to 91.70): V

Ion 134.00 (133.70 to 134.70): V



#84

1,2,4-Trimethylbenzene

Concen: 20.369 ug/l


RT: 13.27 min Scan# 1930

Delta R.T. 0.00 min

Lab File: VD064616.D

Acq: 27 Dec 2019 00:37



Tgt Ion:	105	Resp:	294874
Ion Ratio	Lower	Upper	
105	100		
120	46.8	23.8	71.2




Abundance

Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V









#85
 sec-Butylbenzene
 Concen: 19.386 ug/l
 RT: 13.41 min Scan# 1953
 Delta R.T. 0.00 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1226SBS02

Tot Ion	Ratio	Lower	Upper
105	100		
134	20.9	10.6	31.8




Manual Integrations
 APPROVED


apatel
 12/27/2019 2:07:21 PM



#86
 p-Isopropyltoluene
 Concen: 19.315 ug/l
 RT: 13.52 min Scan# 1972
 Delta R.T. -0.01 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Lower	Upper
119	100		
134	26.9	13.4	40.1
91	22.0	11.2	33.6









#87
 1,3-Dichlorobenzene
 Concen: 20.011 ug/l
 RT: 13.52 min Scan# 1972
 Delta R.T. -0.01 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1226SBSD02

Tot Ion	Ratio	Resp	Lower	Upper
146	100	165861		
111	37.8	18.8	56.4	
148	63.9	32.5	97.5	




Manual Integrations
 APPROVED


apatel
 12/27/2019 2:07:21 PM



#88
 1,4-Dichlorobenzene
 Concen: 20.811 ug/l
 RT: 13.60 min Scan# 1986
 Delta R.T. 0.00 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Resp	Lower	Upper
146	100	169451		
111	35.9	18.1	54.1	
148	63.4	32.3	96.9	









#89
n-Butylbenzene
Concen: 19.064 ug/l
RT: 13.85 min Scan# 2028
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBS02

Tot Ion	Ratio	Lower	Upper
91	100		
92	53.7	26.8	80.4
134	27.9	14.0	41.9




Manual Integrations
APPROVED


apatel
12/27/2019 2:07:21 PM



#90
Hexachloroethane
Concen: 20.159 ug/l
RT: 14.12 min Scan# 2074
Delta R.T. 0.00 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Lower	Upper
117	100		
201	91.0	44.0	131.8









#91
 1,2-Dichlorobenzene
 Concen: 20.803 ug/l
 RT: 13.90 min Scan# 2036
 Delta R.T. 0.00 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1226SBSD02

Tot Ion	Ratio	Lower	Upper
146	100		
111	37.8	19.3	57.9
148	64.7	31.9	95.5




Manual Integrations
 APPROVED

apatel
 12/27/2019 2:07:21 PM



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 21.413 ug/l
 RT: 14.51 min Scan# 2141
 Delta R.T. -0.01 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Lower	Upper
75	100		
155	87.4	48.0	144.0
157	116.9	64.1	192.3









#93
 1,2,4-Trichlorobenzene
 Concen: 19.731 ug/l
 RT: 15.17 min Scan# 2253
 Delta R.T. 0.00 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1226SBSD02

Tot Ion	Ratio	Resp	Lower	Upper
180	100	102679		
182	95.4	48.0	144.2	
145	29.4	13.8	41.4	




Manual Integrations
 APPROVED

apatel
 12/27/2019 2:07:21 PM



#94
 Hexachlorobutadiene
 Concen: 19.665 ug/l
 RT: 15.28 min Scan# 2271
 Delta R.T. 0.00 min
 Lab File: VD064616.D
 Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Resp	Lower	Upper
225	100	64397		
223	58.5	31.1	93.2	
227	60.7	30.9	92.7	









#95
Naphthalene
Concen: 20.081 ug/l
RT: 15.41 min Scan# 2293
Delta R.T. -0.01 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Instrument :
MSVOA_D
ClientSampleId :
VD1226SBS02

Tot Ion	Ratio	Lower	Upper
128	100		
127	13.4	10.3	15.5
129	10.5	8.5	12.7

Manual Integrations
APPROVED

apatel
12/27/2019 2:07:21 PM



#96
1,2,3-Trichlorobenzene
Concen: 20.331 ug/l
RT: 15.60 min Scan# 2326
Delta R.T. -0.01 min
Lab File: VD064616.D
Acq: 27 Dec 2019 00:37

Tgt Ion	Ratio	Lower	Upper
180	100		
182	96.9	47.9	143.7
145	29.0	14.9	44.7

