

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD122719\
 Data File : VD064645.D
 Acq On : 27 Dec 2019 15:23
 Operator : VA/SY
 Sample : K6431-06
 Misc : 5.33G/5.00ml/MSVOA D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-3-(0-2)

Quant Time: Dec 30 03:26:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	502719	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	754811	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	678576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	320433	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	238012	53.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.38%	
35) Dibromofluoromethane	7.92	113	239971	51.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.64%	
50) Toluene-d8	10.34	98	890792	50.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.46%	
62) 4-Bromofluorobenzene	12.64	95	286862	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
Target Compounds						
16) Acetone	4.17	43	8652	12.356	ug/l	93
30) Chloroform	7.71	83	12117	1.399	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD122719\
Data File : VD064645.D
Acq On : 27 Dec 2019 15:23
Operator : VA/SY
Sample : K6431-06
Misc : 5.33G/5.00ml/MSVOA D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-3-(0-2)

Quant Time: Dec 30 03:26:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

