

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122721\
 Data File : VD071479.D
 Acq On : 27 Dec 2021 11:40
 Operator : VA/SY
 Sample : VD1227SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1227SBL01

Quant Time: Dec 27 16:46:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	366919	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	628672	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	555013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	252085	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	243171	49.902	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.800%
35) Dibromofluoromethane	7.914	113	224907	53.230	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.460%
50) Toluene-d8	10.338	98	823453	51.572	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.140%
62) 4-Bromofluorobenzene	12.626	95	287471	50.666	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122721\
Data File : VD071479.D
Acq On : 27 Dec 2021 11:40
Operator : VA/SY
Sample : VD1227SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1227SBL01

Quant Time: Dec 27 16:46:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

