

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122721\
 Data File : VD071491.D
 Acq On : 27 Dec 2021 19:14
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 28 02:45:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	104	0.00
2 T	Dichlorodifluoromethane	0.524	0.443	15.5	94	0.00
3 P	Chloromethane	0.671	0.543	19.1	89	0.00
4 C	Vinyl Chloride	0.804	0.644	19.9#	89	0.00
5 T	Bromomethane	0.562	0.454	19.2	94	0.02
6 T	Chloroethane	0.534	0.438	18.0	91	0.01
7 T	Trichlorofluoromethane	0.984	0.936	4.9	104	0.01
8 T	Diethyl Ether	0.282	0.270	4.3	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.573	0.546	4.7	107	0.01
10 T	Methyl Iodide	0.557	0.601	-7.9	104	0.00
11 T	Tert butyl alcohol	0.065	0.040	38.5#	90	0.00
12 CM	1,1-Dichloroethene	0.521	0.497	4.6#	105	0.01
13 T	Acrolein	0.032	0.026	18.8	84	0.00
14 T	Allyl chloride	0.985	0.965	2.0	106	0.02
15 T	Acrylonitrile	0.139	0.136	2.2	101	0.00
16 T	Acetone	0.167	0.169	-1.2	100	0.00
17 T	Carbon Disulfide	1.697	1.352	20.3	89	0.00
18 T	Methyl Acetate	0.515	0.483	6.2	105	-0.01
19 T	Methyl tert-butyl Ether	1.345	1.350	-0.4	105	0.00
20 T	Methylene Chloride	0.811	0.648	20.1	107	0.00
21 T	trans-1,2-Dichloroethene	0.603	0.569	5.6	104	0.00
22 T	Diisopropyl ether	2.006	2.008	-0.1	106	0.00
23 T	Vinyl Acetate	1.034	1.049	-1.5	102	0.00
24 P	1,1-Dichloroethane	1.150	1.123	2.3	107	0.00
25 T	2-Butanone	0.199	0.192	3.5	98	0.00
26 T	2,2-Dichloropropane	0.945	0.976	-3.3	110	0.00
27 T	cis-1,2-Dichloroethene	0.667	0.658	1.3	105	0.00
28 T	Bromochloromethane	0.471	0.429	8.9	93	0.00
29 T	Tetrahydrofuran	0.120	0.118	1.7	102	0.00
30 C	Chloroform	1.160	1.162	-0.2#	108	0.00
31 T	Cyclohexane	1.128	0.970	14.0	99	0.00
32 T	1,1,1-Trichloroethane	1.029	1.016	1.3	107	0.00
33 S	1,2-Dichloroethane-d4	0.664	0.645	2.9	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.336	0.348	-3.6	104	0.00
36 T	1,1-Dichloropropene	0.537	0.508	5.4	103	0.00
37 T	Ethyl Acetate	0.248	0.240	3.2	97	0.00
38 T	Carbon Tetrachloride	0.530	0.520	1.9	104	0.00
39 T	Methylcyclohexane	0.628	0.578	8.0	100	0.00
40 TM	Benzene	1.443	1.389	3.7	104	0.00
41 T	Methacrylonitrile	0.147	0.146	0.7	92	0.00
42 TM	1,2-Dichloroethane	0.471	0.460	2.3	104	0.00
43 T	Isopropyl Acetate	0.492	0.486	1.2	104	0.00
44 TM	Trichloroethene	0.376	0.371	1.3	108	0.00
45 C	1,2-Dichloropropane	0.375	0.377	-0.5#	109	0.00
46 T	Dibromomethane	0.207	0.202	2.4	102	0.00
47 T	Bromodichloromethane	0.525	0.525	0.0	106	0.00
48 T	Methyl methacrylate	0.243	0.246	-1.2	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122721\
 Data File : VD071491.D
 Acq On : 27 Dec 2021 19:14
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 28 02:45:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	94	0.00
50 S	Toluene-d8	1.270	1.260	0.8	100	0.00
51 T	4-Methyl-2-Pentanone	0.259	0.252	2.7	101	0.00
52 CM	Toluene	0.917	0.900	1.9#	106	0.00
53 T	t-1,3-Dichloropropene	0.491	0.494	-0.6	106	0.00
54 T	cis-1,3-Dichloropropene	0.567	0.562	0.9	106	0.00
55 T	1,1,2-Trichloroethane	0.263	0.270	-2.7	109	0.00
56 T	Ethyl methacrylate	0.346	0.362	-4.6	105	0.00
57 T	1,3-Dichloropropane	0.481	0.486	-1.0	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.166	0.162	2.4	102	0.00
59 T	2-Hexanone	0.183	0.184	-0.5	100	0.00
60 T	Dibromochloromethane	0.330	0.349	-5.8	110	0.00
61 T	1,2-Dibromoethane	0.260	0.260	0.0	105	0.00
62 S	4-Bromofluorobenzene	0.451	0.462	-2.4	105	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
64 T	Tetrachloroethene	0.347	0.327	5.8	106	0.00
65 PM	Chlorobenzene	1.026	1.009	1.7	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.375	0.385	-2.7	108	0.00
67 C	Ethyl Benzene	1.909	1.903	0.3#	106	0.00
68 T	m/p-Xylenes	0.725	0.723	0.3	105	0.00
69 T	o-Xylene	0.663	0.681	-2.7	107	0.00
70 T	Styrene	1.152	1.178	-2.3	107	0.00
71 P	Bromoform	0.203	0.206	-1.5	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.855	3.845	0.3	108	0.00
74 T	N-ethyl acetate	1.034	1.037	-0.3	102	0.00
75 P	1,1,2,2-Tetrachloroethane	0.697	0.690	1.0	105	0.00
76 T	1,2,3-Trichloropropane	0.594	0.482	18.9	87	0.00
77 T	Bromobenzene	0.858	0.829	3.4	105	0.00
78 T	n-propylbenzene	4.832	4.801	0.6	106	0.00
79 T	2-Chlorotoluene	2.700	2.656	1.6	106	0.00
80 T	1,3,5-Trimethylbenzene	3.180	3.227	-1.5	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.210	0.200	4.8	101	0.00
82 T	4-Chlorotoluene	2.801	2.777	0.9	106	0.00
83 T	tert-Butylbenzene	2.717	2.784	-2.5	108	0.00
84 T	1,2,4-Trimethylbenzene	3.147	3.216	-2.2	109	0.00
85 T	sec-Butylbenzene	4.142	4.203	-1.5	108	0.00
86 T	p-Isopropyltoluene	3.437	3.526	-2.6	109	0.00
87 T	1,3-Dichlorobenzene	1.727	1.682	2.6	106	0.00
88 T	1,4-Dichlorobenzene	1.694	1.658	2.1	107	0.00
89 T	n-Butylbenzene	3.373	3.451	-2.3	109	0.00
90 T	Hexachloroethane	0.651	0.650	0.2	107	0.00
91 T	1,2-Dichlorobenzene	1.466	1.448	1.2	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.121	0.119	1.7	111	0.00
93 T	1,2,4-Trichlorobenzene	0.893	0.907	-1.6	107	0.00
94 T	Hexachlorobutadiene	0.479	0.484	-1.0	108	0.00
95 T	Naphthalene	1.701	1.751	-2.9	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122721\
Data File : VD071491.D
Acq On : 27 Dec 2021 19:14
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Dec 28 02:45:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.773	0.786	-1.7	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6