



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122821\
 Data File : VD071492.D
 Acq On : 28 Dec 2021 09:10
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Title : SW846 8260
 Last Update : Thu Dec 23 08:20:24 2021



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1807

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	74147	PASS
75	95	30	60	55.2	182272	PASS
95	95	100	100	100.0	329984	PASS
96	95	5	9	6.4	21221	PASS
173	174	0.00	2	1.7	4255	PASS
174	95	50	100	76.0	250752	PASS
175	174	5	9	7.4	18635	PASS
176	174	95	101	96.1	240853	PASS
177	176	5	9	6.9	16696	PASS