

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122921\
 Data File : VD071525.D
 Acq On : 29 Dec 2021 13:33
 Operator : VA/SY
 Sample : M5230-07
 Misc : 3.93G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PSC164718

Quant Time: Dec 29 15:28:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	348463	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	609759	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	583004	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	264598	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	216654	46.815	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.620%
35) Dibromofluoromethane	7.909	113	194800	47.535	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.060%
50) Toluene-d8	10.332	98	767952	49.588	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.180%
62) 4-Bromofluorobenzene	12.626	95	269934	49.051	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.100%
Target Compounds						
17) Carbon Disulfide	4.409	76	21259	1.797	ug/l	Qvalue # 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122921\
Data File : VD071525.D
Acq On : 29 Dec 2021 13:33
Operator : VA/SY
Sample : M5230-07
Misc : 3.93G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PSC164718

Quant Time: Dec 29 15:28:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

