

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122921\
 Data File : VD071528.D
 Acq On : 29 Dec 2021 14:57
 Operator : VA/SY
 Sample : M5223-05
 Misc : 7.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DB-17-(15-20)-12-22-21

Quant Time: Dec 29 15:30:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	379221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	654519	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	628764	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	292428	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	214249	42.540	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.080%
35) Dibromofluoromethane	7.908	113	208066	47.300	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.600%
50) Toluene-d8	10.332	98	831919	50.045	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.080%
62) 4-Bromofluorobenzene	12.626	95	290561	49.188	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122921\
Data File : VD071528.D
Acq On : 29 Dec 2021 14:57
Operator : VA/SY
Sample : M5223-05
Misc : 7.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DB-17-(15-20)-12-22-21

Quant Time: Dec 29 15:30:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

