



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122921\
Data File : VD071517.D
Acq On : 29 Dec 2021 08:52
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122215.M
Title : SW846 8260
Last Update : Thu Dec 23 08:20:24 2021



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1809

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.6	70920	PASS
75	95	30	60	58.3	175403	PASS
95	95	100	100	100.0	301099	PASS
96	95	5	9	6.7	20071	PASS
173	174	0.00	2	1.0	2284	PASS
174	95	50	100	74.1	223147	PASS
175	174	5	9	7.7	17177	PASS
176	174	95	101	95.8	213675	PASS
177	176	5	9	7.8	16586	PASS