

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD123019\
 Data File : VD064681.D
 Acq On : 30 Dec 2019 17:04
 Operator : VA/SY
 Sample : K6441-03RE
 Misc : 5.78G/5.00ml/MSVOA D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 K321-SB-03RE

Quant Time: Dec 31 08:07:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	36404	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	55540	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	52498	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	26939	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	24187	75.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	150.70%	
35) Dibromofluoromethane	7.92	113	21245	62.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	124.70%	
50) Toluene-d8	10.34	98	62901	48.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.40%	
62) 4-Bromofluorobenzene	12.64	95	22315	54.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.16%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD123019\
Data File : VD064681.D
Acq On : 30 Dec 2019 17:04
Operator : VA/SY
Sample : K6441-03RE
Misc : 5.78G/5.00ml/MSVOA D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
K321-SB-03RE

Quant Time: Dec 31 08:07:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.98 min Scan# 1031

Delta R.T. 0.00 min

Lab File: VD064681.D

Acq: 30 Dec 2019 17:04

Instrument :

MSVOA_D

ClientSampleId :


K321-SB-03RE

Tot Ion:168 Resp: 36404

Ion Ratio Lower Upper

168 100

99 50.0 41.8 62.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC

7.98

15000



10000

5000

0

7.90 7.95 8.00 8.05

Time-->



#33

1,2-Dichloroethane-d4

Concen: 75.347 ug/l

RT: 8.34 min Scan# 1091

Delta R.T. 0.01 min

Lab File: VD064681.D


Acq: 30 Dec 2019 17:04

Tgt Ion: 65 Resp: 24187

Ion Ratio Lower Upper

65 100

67 53.6 0.0 106.2



Abundance Ion 64.90 (64.60 to 65.60): VDC

Ion 66.90 (66.60 to 67.60): VDC

8.34

12000

10000

8000

6000


4000


2000

0

8.25 8.30 8.35 8.40

Time-->









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.87 min Scan# 1181
 Delta R.T. 0.00 min
 Lab File: VD064681.D
 Acq: 30 Dec 2019 17:04

Instrument :
 MSVOA_D
 ClientSampleId :
 K321-SB-03RE

Tot Ion	Ratio	Lower	Upper
114	100		
63	21.4	0.0	38.0
88	12.7	0.0	29.2




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): V
 Ion 87.90 (87.60 to 88.60): V






#35
 Dibromofluoromethane
 Concen: 62.348 ug/l
 RT: 7.92 min Scan# 1020
 Delta R.T. 0.00 min
 Lab File: VD064681.D
 Acq: 30 Dec 2019 17:04

Tgt Ion	Ratio	Lower	Upper
113	100		
111	98.9	80.6	121.0
192	21.5	17.9	26.9



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.199 ug/l

RT: 10.34 min Scan# 1432

Delta R.T. 0.01 min

Lab File: VD064681.D

Acq: 30 Dec 2019 17:04

Instrument :

MSVOA_D

ClientSampleId :





K321-SB-03RE

Tot Ion: 98 Resp: 62901

Ion Ratio Lower Upper

98 100

100 66.8 51.7 77.5



#62

4-Bromofluorobenzene

Concen: 54.076 ug/l

RT: 12.64 min Scan# 1822

Delta R.T. 0.00 min

Lab File: VD064681.D

Acq: 30 Dec 2019 17:04




Tgt Ion: 95 Resp: 22315


Ion Ratio Lower Upper

95 100

174 93.4 0.0 193.2

176 87.0 0.0 184.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.65 min Scan# 1654
Delta R.T. 0.00 min
Lab File: VD064681.D
Acq: 30 Dec 2019 17:04

Instrument :
MSVOA_D
ClientSampleId :
K321-SB-03RE

Tot Ion:117 Resp: 52498
Ion Ratio Lower Upper
117 100
82 59.3 42.2 63.4
119 35.8 26.0 39.0




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.58 min Scan# 1983
Delta R.T. 0.00 min
Lab File: VD064681.D
Acq: 30 Dec 2019 17:04

Tgt Ion:152 Resp: 26939
Ion Ratio Lower Upper
152 100
115 53.8 27.8 83.4
150 149.2 0.0 351.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

