

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD123021\
 Data File : VD071541.D
 Acq On : 30 Dec 2021 11:30
 Operator : VA/SY
 Sample : M5210-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 124027

Quant Time: Dec 30 15:30:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.973	168	121453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	210064	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	221682	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	106965	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	108819	67.464	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 134.920%	
35) Dibromofluoromethane	7.908	113	76941	54.499	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.000%	
50) Toluene-d8	10.338	98	268606	50.346	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.700%	
62) 4-Bromofluorobenzene	12.626	95	102660	54.150	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 108.300%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.173	59	3445	Below Cal	#	71
16) Acetone	4.162	43	6910	17.055	ug/l #	75
20) Methylene Chloride	4.956	84	3898	Below Cal		93
25) 2-Butanone	7.197	43	1491	3.081	ug/l	100
31) Cyclohexane	7.973	56	3427	1.251	ug/l #	26
36) 1,1-Dichloropropene	7.973	75	11959	5.302	ug/l #	47
38) Carbon Tetrachloride	7.973	117	14068	6.322	ug/l #	17
76) 1,2,3-Trichloropropane	12.626	75	60919	47.932	ug/l #	1
81) trans-1,4-Dichloro-2-b...	12.626	75	60919	135.434	ug/l #	18


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD123021\
Data File : VD071541.D
Acq On : 30 Dec 2021 11:30
Operator : VA/SY
Sample : M5210-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
124027

Quant Time: Dec 30 15:30:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.973 min Scan# 1030
Delta R.T. 0.000 min
Lab File: VD071541.D
Acq: 30 Dec 2021 11:30




Instrument :
MSVOA_D
ClientSampleId :
124027


Tgt Ion:168 Resp: 121453
Ion Ratio Lower Upper
168 100
99 67.3 46.6 69.8



#16
Acetone
Concen: 17.055 ug/l
RT: 4.162 min Scan# 381
Delta R.T. 0.006 min
Lab File: VD071541.D
Acq: 30 Dec 2021 11:30

Tgt Ion: 43 Resp: 6910
Ion Ratio Lower Upper
43 100
58 20.3 27.8 41.8#









#20
Methylene Chloride
Concen: Below Cal
RT: 4.956 min Scan# 516
Delta R.T. -0.005 min
Lab File: VD071541.D
Acq: 30 Dec 2021 11:30




Instrument :
MSVOA_D
ClientSampleId :
124027


Tgt Ion:	84	Resp:	3898
Ion	Ratio	Lower	Upper
84	100		
49	107.8	95.8	143.6
51	39.5	30.9	46.3
86	60.7	50.3	75.5



#25
2-Butanone
Concen: 3.081 ug/l
RT: 7.197 min Scan# 897
Delta R.T. -0.011 min
Lab File: VD071541.D
Acq: 30 Dec 2021 11:30

Tgt Ion:	43	Resp:	1491
Ion	Ratio	Lower	Upper
43	100		
72	29.5	23.8	35.6





#31

Cyclohexane

Concen: 1.251 ug/l

RT: 7.973 min Scan# 1029

Delta R.T. -0.012 min

Lab File: VD071541.D

Acq: 30 Dec 2021 11:30

Instrument :

MSVOA_D

ClientSampleId :

124027





Tgt Ion: 56 Resp: 3427

Ion Ratio Lower Upper

56 100

69 133.5 25.9 38.9#

84 108.5 64.3 96.5#



#33

1,2-Dichloroethane-d4

Concen: 67.464 ug/l

RT: 8.332 min Scan# 1090

Delta R.T. 0.006 min

Lab File: VD071541.D




Acq: 30 Dec 2021 11:30


Tgt Ion: 65 Resp: 108819

Ion Ratio Lower Upper

65 100

67 49.0 0.0 112.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.861 min Scan# 1180

Delta R.T. 0.000 min

Lab File: VD071541.D

Acq: 30 Dec 2021 11:30

Instrument :

MSVOA_D

ClientSampleId :

124027





Tgt Ion:114 Resp: 210064

Ion Ratio Lower Upper

114 100

63 22.3 0.0 41.8

88 15.2 0.0 32.8



#35

Dibromofluoromethane

Concen: 54.499 ug/l

RT: 7.908 min Scan# 1018

Delta R.T. 0.000 min

Lab File: VD071541.D

Acq: 30 Dec 2021 11:30




Tgt Ion:113 Resp: 76941


Ion Ratio Lower Upper

113 100

111 101.3 82.0 123.0

192 16.4 15.2 22.8









#36
1,1-Dichloropropene
Concen: 5.302 ug/l
RT: 7.973 min Scan# 1029
Delta R.T. -0.135 min
Lab File: VD071541.D
Acq: 30 Dec 2021 11:30




Instrument :
MSVOA_D
ClientSampleId :
124027


Tgt Ion: 75 Resp: 11959
Ion Ratio Lower Upper
75 100
110 5.9 17.8 53.4#
77 0.0 24.8 37.2#



#38
Carbon Tetrachloride
Concen: 6.322 ug/l
RT: 7.973 min Scan# 1029
Delta R.T. -0.124 min
Lab File: VD071541.D
Acq: 30 Dec 2021 11:30

Tgt Ion: 117 Resp: 14068
Ion Ratio Lower Upper
117 100
119 6.1 75.8 113.6#
121 0.0 23.6 35.4#









#50
Toluene-d8
Concen: 50.346 ug/l
RT: 10.338 min Scan# 1430
Delta R.T. 0.006 min
Lab File: VD071541.D
Acq: 30 Dec 2021 11:30




Instrument :
MSVOA_D
ClientSampleId :
124027


Tgt Ion: 98 Resp: 268606
Ion Ratio Lower Upper
98 100
100 64.9 51.8 77.6



#62
4-Bromofluorobenzene
Concen: 54.150 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. -0.000 min
Lab File: VD071541.D
Acq: 30 Dec 2021 11:30

Tgt Ion: 95 Resp: 102660
Ion Ratio Lower Upper
95 100
174 77.2 0.0 159.4
176 74.8 0.0 154.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 1652
Delta R.T. -0.000 min
Lab File: VD071541.D
Acq: 30 Dec 2021 11:30




Instrument :
MSVOA_D
ClientSampleId :
124027


Tgt Ion:117 Resp: 221682
Ion Ratio Lower Upper
117 100
82 58.6 44.7 67.1
119 32.4 26.2 39.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VD071541.D
Acq: 30 Dec 2021 11:30

Tgt Ion:152 Resp: 106965
Ion Ratio Lower Upper
152 100
115 61.1 29.8 89.3
150 157.6 0.0 349.2









#76
1,2,3-Trichloropropane
Concen: 47.932 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. -0.147 min
Lab File: VD071541.D
Acq: 30 Dec 2021 11:30

Instrument :
MSVOA_D
ClientSampleId :
124027

Tgt Ion: 75 Resp: 60919
Ion Ratio Lower Upper
75 100
77 1.0 158.4 475.3#



#81
trans-1,4-Dichloro-2-butene
Concen: 135.434 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. 0.106 min
Lab File: VD071541.D
Acq: 30 Dec 2021 11:30

Tgt Ion: 75 Resp: 60919
Ion Ratio Lower Upper
75 100
53 0.1 62.6 94.0#
89 0.0 34.5 51.7#

