

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD123021\
 Data File : VD071538.D
 Acq On : 30 Dec 2021 10:06
 Operator : VA/SY
 Sample : VD1230SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1230SBL01

Quant Time: Dec 30 15:27:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	274458	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	485614	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	446379	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	200837	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	195301	53.580	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.160%
35) Dibromofluoromethane	7.914	113	165545	50.723	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.440%
50) Toluene-d8	10.338	98	586122	47.522	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.040%
62) 4-Bromofluorobenzene	12.626	95	199746	45.576	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD123021\
Data File : VD071538.D
Acq On : 30 Dec 2021 10:06
Operator : VA/SY
Sample : VD1230SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1230SBL01

Quant Time: Dec 30 15:27:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

