

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD123021\
 Data File : VD071546.D
 Acq On : 30 Dec 2021 13:50
 Operator : VA/SY
 Sample : M5239-02
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SASP2-123-1-5-10

Quant Time: Dec 30 15:33:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	450085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	761103	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	739225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	342354	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	248986	41.654	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	83.300%
35) Dibromofluoromethane	7.914	113	238675	46.660	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.320%
50) Toluene-d8	10.338	98	956999	49.507	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.020%
62) 4-Bromofluorobenzene	12.620	95	347851	50.640	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD123021\
Data File : VD071546.D
Acq On : 30 Dec 2021 13:50
Operator : VA/SY
Sample : M5239-02
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SASP2-123-1-5-10

Quant Time: Dec 30 15:33:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

