

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD123021\
 Data File : VD071552.D
 Acq On : 30 Dec 2021 16:38
 Operator : VA/SY
 Sample : M5239-03
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SASP2-123-1-10-37

Quant Time: Dec 30 17:23:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	341200	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	581972	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	566241	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	273554	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	208336	45.976	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.960%
35) Dibromofluoromethane	7.909	113	185805	47.505	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.000%
50) Toluene-d8	10.332	98	730457	49.419	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.840%
62) 4-Bromofluorobenzene	12.620	95	271077	51.610	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD123021\
Data File : VD071552.D
Acq On : 30 Dec 2021 16:38
Operator : VA/SY
Sample : M5239-03
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SASP2-123-1-10-37

Quant Time: Dec 30 17:23:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

