

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD123021\
 Data File : VD071557.D
 Acq On : 30 Dec 2021 18:59
 Operator : VA/SY
 Sample : M5337-09
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-1

Quant Time: Dec 31 06:57:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	25167	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	44047	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	50714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	29618	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	22561	67.499	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.000%
35) Dibromofluoromethane	7.914	113	13517	45.661	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.320%
50) Toluene-d8	10.332	98	59438	53.131	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.260%
62) 4-Bromofluorobenzene	12.626	95	28648	72.065	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	144.120%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD123021\
Data File : VD071557.D
Acq On : 30 Dec 2021 18:59
Operator : VA/SY
Sample : M5337-09
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-1

Quant Time: Dec 31 06:57:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

