

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD123021\
 Data File : VD071543.D
 Acq On : 30 Dec 2021 12:26
 Operator : VA/SY
 Sample : M5237-01
 Misc : 4.21G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-1S

Quant Time: Dec 30 15:31:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	378944	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	641468	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	624825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	284292	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	217453	43.208	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	86.420%
35) Dibromofluoromethane	7.914	113	201347	46.704	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.400%
50) Toluene-d8	10.338	98	822456	50.482	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.960%
62) 4-Bromofluorobenzene	12.626	95	283591	48.985	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.980%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD123021\
Data File : VD071543.D
Acq On : 30 Dec 2021 12:26
Operator : VA/SY
Sample : M5237-01
Misc : 4.21G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-1S

Quant Time: Dec 30 15:31:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.973 min Scan# 1030
Delta R.T. 0.000 min
Lab File: VD071543.D
Acq: 30 Dec 2021 12:26




Instrument :
MSVOA_D
ClientSampleId :
TP-1S


Tgt Ion:168 Resp: 378944
Ion Ratio Lower Upper
168 100
99 59.2 46.6 69.8



#33
1,2-Dichloroethane-d4
Concen: 43.208 ug/l
RT: 8.326 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VD071543.D
Acq: 30 Dec 2021 12:26

Tgt Ion: 65 Resp: 217453
Ion Ratio Lower Upper
65 100
67 50.9 0.0 112.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.861 min Scan# 1180

Delta R.T. 0.000 min

Lab File: VD071543.D

Acq: 30 Dec 2021 12:26

Instrument :

MSVOA_D

ClientSampleId :

TP-1S





Tgt Ion:114 Resp: 641468

Ion Ratio Lower Upper

114 100

63 25.1 0.0 41.8

88 15.3 0.0 32.8



#35

Dibromofluoromethane

Concen: 46.704 ug/l

RT: 7.914 min Scan# 1019

Delta R.T. 0.006 min

Lab File: VD071543.D

Acq: 30 Dec 2021 12:26




Tgt Ion:113 Resp: 201347


Ion Ratio Lower Upper

113 100

111 104.5 82.0 123.0

192 17.6 15.2 22.8









#50
Toluene-d8
Concen: 50.482 ug/l
RT: 10.338 min Scan# 1430
Delta R.T. 0.006 min
Lab File: VD071543.D
Acq: 30 Dec 2021 12:26




Instrument :
MSVOA_D
ClientSampleId :
TP-1S


Tgt Ion: 98 Resp: 822456
Ion Ratio Lower Upper
98 100
100 64.8 51.8 77.6



#62
4-Bromofluorobenzene
Concen: 48.985 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. -0.000 min
Lab File: VD071543.D
Acq: 30 Dec 2021 12:26

Tgt Ion: 95 Resp: 283591
Ion Ratio Lower Upper
95 100
174 76.1 0.0 159.4
176 73.6 0.0 154.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD071543.D
Acq: 30 Dec 2021 12:26

Instrument :
MSVOA_D
ClientSampleId :
TP-1S

Tgt Ion:117 Resp: 624825
Ion Ratio Lower Upper
117 100
82 59.6 44.7 67.1
119 31.3 26.2 39.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.567 min Scan# 1980
Delta R.T. 0.006 min
Lab File: VD071543.D
Acq: 30 Dec 2021 12:26

Tgt Ion:152 Resp: 284292
Ion Ratio Lower Upper
152 100
115 60.8 29.8 89.3
150 157.7 0.0 349.2

