

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD123021\
 Data File : VD071553.D
 Acq On : 30 Dec 2021 17:06
 Operator : VA/SY
 Sample : M5210-01RE
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 124027RE

Quant Time: Dec 30 17:24:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	19418	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	33870	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	35496	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	17425	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	16905	65.552	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.100%
35) Dibromofluoromethane	7.902	113	10878	47.787	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.580%
50) Toluene-d8	10.332	98	39936	46.425	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.840%
62) 4-Bromofluorobenzene	12.620	95	16321	53.392	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.780%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD123021\
Data File : VD071553.D
Acq On : 30 Dec 2021 17:06
Operator : VA/SY
Sample : M5210-01RE
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
124027RE

Quant Time: Dec 30 17:24:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.973 min Scan# 1029
Delta R.T. 0.000 min
Lab File: VD071553.D
Acq: 30 Dec 2021 17:06




Instrument :
MSVOA_D
ClientSampleId :
124027RE

Tgt Ion:168 Resp: 19418
Ion Ratio Lower Upper
168 100
99 61.5 46.6 69.8



#33
1,2-Dichloroethane-d4
Concen: 65.552 ug/l
RT: 8.326 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VD071553.D
Acq: 30 Dec 2021 17:06

Tgt Ion: 65 Resp: 16905
Ion Ratio Lower Upper
65 100
67 53.3 0.0 112.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.861 min Scan# 1180

Delta R.T. 0.000 min

Lab File: VD071553.D

Acq: 30 Dec 2021 17:06

Instrument :

MSVOA_D

ClientSampleId :

124027RE





Tgt Ion:114 Resp: 33870

Ion Ratio Lower Upper

114 100

63 22.6 0.0 41.8

88 11.6 0.0 32.8



#35

Dibromofluoromethane

Concen: 47.787 ug/l

RT: 7.902 min Scan# 1017

Delta R.T. -0.006 min

Lab File: VD071553.D

Acq: 30 Dec 2021 17:06




Tgt Ion:113 Resp: 10878


Ion Ratio Lower Upper

113 100

111 101.1 82.0 123.0

192 16.3 15.2 22.8









#50
Toluene-d8
Concen: 46.425 ug/l
RT: 10.332 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VD071553.D
Acq: 30 Dec 2021 17:06




Instrument :
MSVOA_D
ClientSampleId :
124027RE


Tgt Ion: 98 Resp: 39936
Ion Ratio Lower Upper
98 100
100 64.1 51.8 77.6



#62
4-Bromofluorobenzene
Concen: 53.392 ug/l
RT: 12.620 min Scan# 1819
Delta R.T. -0.006 min
Lab File: VD071553.D
Acq: 30 Dec 2021 17:06

Tgt Ion: 95 Resp: 16321
Ion Ratio Lower Upper
95 100
174 77.5 0.0 159.4
176 79.6 0.0 154.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 117
Delta R.T. -0.000 min
Lab File: VD071553.D
Acq: 30 Dec 2021 17:06

Instrument :
MSVOA_D
ClientSampleId :
124027RE

Tgt Ion:117 Resp: 35496
Ion Ratio Lower Upper
117 100
82 58.3 44.7 67.1
119 31.9 26.2 39.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VD071553.D
Acq: 30 Dec 2021 17:06

Tgt Ion:152 Resp: 17425
Ion Ratio Lower Upper
152 100
115 63.1 29.8 89.3
150 158.0 0.0 349.2

