

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD123119\
 Data File : VD064689.D
 Acq On : 31 Dec 2019 12:24
 Operator : VA/SY
 Sample : K6450-03
 Misc : 7.60G/5.00ml/MSVOA D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WP022SB448_191226-30-32

Quant Time: Jan 01 01:22:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	483122	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	746792	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	608642	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	228592	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	211102	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
35) Dibromofluoromethane	7.92	113	227438	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
50) Toluene-d8	10.34	98	881520	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
62) 4-Bromofluorobenzene	12.64	95	236851	42.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.38%	
Target Compounds						
4) Vinyl Chloride	2.35	62	62031	9.689	ug/l	94
16) Acetone	4.16	43	7116	11.176	ug/l	93
30) Chloroform	7.71	83	11493	1.380	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD123119\
Data File : VD064689.D
Acq On : 31 Dec 2019 12:24
Operator : VA/SY
Sample : K6450-03
Misc : 7.60G/5.00ml/MSVOA D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WP022SB448_191226-30-32

Quant Time: Jan 01 01:22:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

