

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD123119\
 Data File : VD064692.D
 Acq On : 31 Dec 2019 13:51
 Operator : VA/SY
 Sample : K6455-03
 Misc : 6.54G/5.00ml/MSVOA D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 K435-SB3

Quant Time: Jan 01 01:29:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	490219	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	768283	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	683490	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	326436	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	242034	55.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.98%	
35) Dibromofluoromethane	7.92	113	242755	51.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.00%	
50) Toluene-d8	10.34	98	925810	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
62) 4-Bromofluorobenzene	12.64	95	282576	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
Target Compounds						
30) Chloroform	7.71	83	11230	1.329	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD123119\
Data File : VD064692.D
Acq On : 31 Dec 2019 13:51
Operator : VA/SY
Sample : K6455-03
Misc : 6.54G/5.00ml/MSVOA D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
K435-SB3

Quant Time: Jan 01 01:29:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

