


Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD123119\
Data File : VD064683.D
Acq On : 31 Dec 2019 08:40
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Title : SW846 8260
Last Update : Tue Dec 24 13:19:06 2019



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.8	34986	PASS
75	95	30	60	54.2	91093	PASS
95	95	100	100	100.0	168130	PASS
96	95	5	9	6.2	10456	PASS
173	174	0.00	2	0.7	1043	PASS
174	95	50	100	95.2	160064	PASS
175	174	5	9	7.6	12208	PASS
176	174	95	101	95.2	152341	PASS
177	176	5	9	7.0	10716	PASS