

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010721\
 Data File : VD068024.D
 Acq On : 07 Jan 2021 10:32
 Operator : VA/SY
 Sample : VD0107SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0107SBL01

Quant Time: Jan 08 00:12:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 23 07:34:16 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	287258	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	569759	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	555046	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	244299	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	198261	63.91	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.82%
35) Dibromofluoromethane	7.914	113	196192	54.65	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.30%
50) Toluene-d8	10.344	98	730969	54.17	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.34%
62) 4-Bromofluorobenzene	12.632	95	231166	50.08	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.16%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010721\
Data File : VD068024.D
Acq On : 07 Jan 2021 10:32
Operator : VA/SY
Sample : VD0107SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0107SBL01

Quant Time: Jan 08 00:12:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 23 07:34:16 2020
Response via : Initial Calibration

