

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010721\
 Data File : VD068032.D
 Acq On : 07 Jan 2021 14:18
 Operator : VA/SY
 Sample : M1044-02
 Misc : 5.19G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-2

Quant Time: Jan 08 00:15:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 23 07:34:16 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	297640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	572884	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	572104	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	251446	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	192987	60.04	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 120.08%	
35) Dibromofluoromethane	7.914	113	188728	52.28	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.56%	
50) Toluene-d8	10.344	98	736746	54.30	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 108.60%	
62) 4-Bromofluorobenzene	12.632	95	249975	53.86	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 107.72%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010721\
Data File : VD068032.D
Acq On : 07 Jan 2021 14:18
Operator : VA/SY
Sample : M1044-02
Misc : 5.19G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-2

Quant Time: Jan 08 00:15:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 23 07:34:16 2020
Response via : Initial Calibration

