

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010721\
 Data File : VD068034.D
 Acq On : 07 Jan 2021 15:12
 Operator : VA/SY
 Sample : M1044-06
 Misc : 5.36G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-6

Manual Integrations
 APPROVED

MMDadoda
 1/8/2021 2:32:04 PM

Quant Time: Jan 08 00:15:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 23 07:34:16 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	288042	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	560418	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	531638	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	232925	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	185240	59.55	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 119.10%			
35) Dibromofluoromethane	7.920	113	185957	52.66	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 105.32%			
50) Toluene-d8	10.344	98	699607	52.71	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 105.42%			
62) 4-Bromofluorobenzene	12.632	95	231803	51.06	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 102.12%			
Target Compounds						Qvalue
16) Acetone	4.162	43	7261m	12.23	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010721\
Data File : VD068034.D
Acq On : 07 Jan 2021 15:12
Operator : VA/SY
Sample : M1044-06
Misc : 5.36G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-6

Quant Time: Jan 08 00:15:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 23 07:34:16 2020
Response via : Initial Calibration

Manual Integrations
APPROVED

MMDadoda
1/8/2021 2:32:04 PM

