

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010721\
 Data File : VD068036.D
 Acq On : 07 Jan 2021 16:07
 Operator : VA/SY
 Sample : M1044-10
 Misc : 5.74G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-10

Quant Time: Jan 08 00:16:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 23 07:34:16 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	289770	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	545134	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	526213	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	225536	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	184297	58.90	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 117.80%			
35) Dibromofluoromethane	7.920	113	181384	52.81	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 105.62%			
50) Toluene-d8	10.344	98	696127	53.92	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 107.84%			
62) 4-Bromofluorobenzene	12.632	95	225492	51.06	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 102.12%			
Target Compounds						Qvalue
16) Acetone	4.168	43	9694	16.23	ug/l	95
68) m/p-Xylenes	11.849	106	22870	2.96	ug/l	93
69) o-Xylene	12.185	106	7538	1.09	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010721\
Data File : VD068036.D
Acq On : 07 Jan 2021 16:07
Operator : VA/SY
Sample : M1044-10
Misc : 5.74G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-10

Quant Time: Jan 08 00:16:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 23 07:34:16 2020
Response via : Initial Calibration

