

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010721\
 Data File : VD068023.D
 Acq On : 07 Jan 2021 09:57
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 08 00:11:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 23 07:34:16 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.509	0.530	-4.1	88	0.00
3 P	Chloromethane	0.696	0.584	16.1	76	0.00
4 C	Vinyl Chloride	0.649	0.596	8.2#	81	0.00
5 T	Bromomethane	0.467	0.404	13.5	82	0.00
6 T	Chloroethane	0.399	0.358	10.3	81	0.00
7 T	Trichlorofluoromethane	0.956	0.987	-3.2	92	0.00
8 T	Diethyl Ether	0.274	0.236	13.9	73	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.570	0.564	1.1	88	0.00
10 T	Methyl Iodide	0.498	0.525	-5.4	77	0.00
11 T	Tert butyl alcohol	0.046	0.024	47.8#	65	-0.01
12 CM	1,1-Dichloroethene	0.547	0.530	3.1#	84	0.00
13 T	Acrolein	0.038	0.036	5.3	79	0.00
14 T	Allyl chloride	0.972	0.848	12.8	73	0.00
15 T	Acrylonitrile	0.131	0.112	14.5	75	0.00
16 T	Acetone	0.103	0.112	-8.7	102	0.00
17 T	Carbon Disulfide	1.905	1.773	6.9	79	0.00
18 T	Methyl Acetate	0.305	0.257	15.7	74	0.00
19 T	Methyl tert-butyl Ether	1.137	1.027	9.7	76	0.00
20 T	Methylene Chloride	0.849	0.617	27.3#	76	0.00
21 T	trans-1,2-Dichloroethene	0.653	0.620	5.1	80	0.00
22 T	Diisopropyl ether	1.958	1.772	9.5	75	0.00
23 T	Vinyl Acetate	0.995	0.893	10.3	73	0.00
24 P	1,1-Dichloroethane	1.185	1.086	8.4	80	0.00
25 T	2-Butanone	0.152	0.132	13.2	76	0.00
26 T	2,2-Dichloropropane	0.934	0.935	-0.1	89	0.00
27 T	cis-1,2-Dichloroethene	0.698	0.657	5.9	80	0.00
28 T	Bromochloromethane	0.447	0.435	2.7	80	0.00
29 T	Tetrahydrofuran	0.103	0.088	14.6	73	-0.01
30 C	Chloroform	1.161	1.114	4.0#	82	0.00
31 T	Cyclohexane	1.104	0.986	10.7	81	0.00
32 T	1,1,1-Trichloroethane	0.994	0.970	2.4	86	0.00
33 S	1,2-Dichloroethane-d4	0.540	0.497	8.0	84	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	76	0.00
35 S	Dibromofluoromethane	0.315	0.328	-4.1	85	0.00
36 T	1,1-Dichloropropene	0.531	0.553	-4.1	83	0.00
37 T	Ethyl Acetate	0.215	0.202	6.0	75	-0.01
38 T	Carbon Tetrachloride	0.508	0.545	-7.3	89	0.00
39 T	Methylcyclohexane	0.609	0.634	-4.1	86	0.00
40 TM	Benzene	1.542	1.552	-0.6	80	0.00
41 T	Methacrylonitrile	0.121	0.103	14.9	67	0.00
42 TM	1,2-Dichloroethane	0.416	0.414	0.5	82	0.00
43 T	Isopropyl Acetate	0.395	0.355	10.1	71	0.00
44 TM	Trichloroethene	0.386	0.395	-2.3	83	0.00
45 C	1,2-Dichloropropane	0.392	0.390	0.5#	80	0.00
46 T	Dibromomethane	0.191	0.197	-3.1	84	0.00
47 T	Bromodichloromethane	0.515	0.530	-2.9	82	0.00
48 T	Methyl methacrylate	0.204	0.196	3.9	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010721\
 Data File : VD068023.D
 Acq On : 07 Jan 2021 09:57
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 08 00:11:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 23 07:34:16 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	74	0.00
50 S	Toluene-d8	1.184	1.188	-0.3	84	0.00
51 T	4-Methyl-2-Pentanone	0.204	0.193	5.4	73	0.00
52 CM	Toluene	0.952	0.988	-3.8#	82	0.00
53 T	t-1,3-Dichloropropene	0.466	0.476	-2.1	81	0.00
54 T	cis-1,3-Dichloropropene	0.576	0.579	-0.5	80	0.00
55 T	1,1,2-Trichloroethane	0.267	0.263	1.5	80	0.00
56 T	Ethyl methacrylate	0.300	0.301	-0.3	77	0.00
57 T	1,3-Dichloropropane	0.469	0.475	-1.3	80	0.00
58 T	2-Chloroethyl Vinyl ether	0.151	0.162	-7.3	77	0.00
59 T	2-Hexanone	0.134	0.142	-6.0	83	0.00
60 T	Dibromochloromethane	0.330	0.337	-2.1	83	0.00
61 T	1,2-Dibromoethane	0.256	0.256	0.0	81	0.00
62 S	4-Bromofluorobenzene	0.405	0.415	-2.5	85	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	76	0.00
64 T	Tetrachloroethene	0.351	0.356	-1.4	86	0.00
65 PM	Chlorobenzene	1.079	1.072	0.6	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.398	-3.4	84	0.00
67 C	Ethyl Benzene	1.924	1.978	-2.8#	82	0.00
68 T	m/p-Xylenes	0.733	0.768	-4.8	83	0.00
69 T	o-Xylene	0.656	0.668	-1.8	80	0.00
70 T	Styrene	1.134	1.171	-3.3	80	0.00
71 P	Bromoform	0.195	0.186	4.6	77	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
73 T	Isopropylbenzene	3.711	3.806	-2.6	85	0.00
74 T	N-amyl acetate	0.842	0.723	14.1	70	0.00
75 P	1,1,2,2-Tetrachloroethane	0.689	0.654	5.1	82	0.00
76 T	1,2,3-Trichloropropane	0.421	0.491	-16.6	94	0.00
77 T	Bromobenzene	0.849	0.841	0.9	81	0.00
78 T	n-propylbenzene	4.613	4.783	-3.7	85	0.00
79 T	2-Chlorotoluene	2.606	2.586	0.8	82	0.00
80 T	1,3,5-Trimethylbenzene	3.082	3.224	-4.6	85	0.00
81 T	trans-1,4-Dichloro-2-butene	0.205	0.202	1.5	83	0.00
82 T	4-Chlorotoluene	2.739	2.785	-1.7	84	0.00
83 T	tert-Butylbenzene	2.535	2.618	-3.3	84	0.00
84 T	1,2,4-Trimethylbenzene	3.091	3.174	-2.7	83	0.00
85 T	sec-Butylbenzene	3.709	3.839	-3.5	86	0.00
86 T	p-Isopropyltoluene	3.264	3.427	-5.0	86	0.00
87 T	1,3-Dichlorobenzene	1.688	1.626	3.7	81	0.00
88 T	1,4-Dichlorobenzene	1.667	1.623	2.6	83	0.00
89 T	n-Butylbenzene	3.238	3.408	-5.3	86	0.00
90 T	Hexachloroethane	0.667	0.673	-0.9	85	0.00
91 T	1,2-Dichlorobenzene	1.421	1.412	0.6	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.108	0.090	16.7	79	0.00
93 T	1,2,4-Trichlorobenzene	0.844	0.831	1.5	83	0.00
94 T	Hexachlorobutadiene	0.494	0.508	-2.8	91	0.00
95 T	Naphthalene	1.491	1.397	6.3	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010721\
Data File : VD068023.D
Acq On : 07 Jan 2021 09:57
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jan 08 00:11:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 23 07:34:16 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.729	0.715	1.9	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6