


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010721\
Data File : VD068022.D
Acq On : 07 Jan 2021 09:22
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
Title : SW846 8260
Last Update : Wed Dec 23 07:34:16 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	35552	PASS
75	95	30	60	51.4	92907	PASS
95	95	100	100	100.0	180907	PASS
96	95	5	9	6.3	11407	PASS
173	174	0.00	2	1.5	2115	PASS
174	95	50	100	76.4	138243	PASS
175	174	5	9	7.6	10488	PASS
176	174	95	101	96.2	133003	PASS
177	176	5	9	6.7	8941	PASS