

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011321\
 Data File : VD068064.D
 Acq On : 13 Jan 2021 12:01
 Operator : VA/SY
 Sample : M1091-01
 Misc : 6.32G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-03-011221

Quant Time: Jan 13 12:26:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011221S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 13 03:57:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	604800	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	1008049	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.650	117	972057	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	445411	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	285347	55.49	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 110.98%	
35) Dibromofluoromethane	7.920	113	311400	54.20	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.40%	
50) Toluene-d8	10.344	98	1172046	55.35	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 110.70%	
62) 4-Bromofluorobenzene	12.638	95	428560	57.76	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 115.52%	
Target Compounds						
						Qvalue
52) Toluene	10.408	92	61881	3.71	ug/l	94
68) m/p-Xylenes	11.855	106	23899	1.85	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011321\
Data File : VD068064.D
Acq On : 13 Jan 2021 12:01
Operator : VA/SY
Sample : M1091-01
Misc : 6.32G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-03-011221

Quant Time: Jan 13 12:26:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011221S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 13 03:57:57 2021
Response via : Initial Calibration

