

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011321\
 Data File : VD068069.D
 Acq On : 13 Jan 2021 14:18
 Operator : VA/SY
 Sample : M1069-02
 Misc : 6.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-GF-02-065-A

Quant Time: Jan 13 15:26:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011221S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 13 03:57:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	533897	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	900298	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	898526	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	432259	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	251442	55.39	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.78%
35) Dibromofluoromethane	7.920	113	282193	54.99	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.98%
50) Toluene-d8	10.344	98	1077192	56.95	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	113.90%
62) 4-Bromofluorobenzene	12.638	95	384879	58.09	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.18%
Target Compounds						
16) Acetone	4.162	43	664750	923.92	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011321\
Data File : VD068069.D
Acq On : 13 Jan 2021 14:18
Operator : VA/SY
Sample : M1069-02
Misc : 6.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-GF-02-065-A

Quant Time: Jan 13 15:26:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011221S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 13 03:57:57 2021
Response via : Initial Calibration

