

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011321\
 Data File : VD068077.D
 Acq On : 13 Jan 2021 18:23
 Operator : VA/SY
 Sample : M1121-01
 Misc : 6.17G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NP-SP

Quant Time: Jan 14 10:36:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011221S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 13 03:57:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	531224	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	896605	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	872225	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	404536	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	258568	57.24	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.48%
35) Dibromofluoromethane	7.920	113	286735	56.11	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.22%
50) Toluene-d8	10.344	98	1079472	57.31	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.62%
62) 4-Bromofluorobenzene	12.638	95	374334	56.73	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.46%
Target Compounds						
16) Acetone	4.167	43	10594	14.80	ug/l	Qvalue # 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011321\
Data File : VD068077.D
Acq On : 13 Jan 2021 18:23
Operator : VA/SY
Sample : M1121-01
Misc : 6.17G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NP-SP

Quant Time: Jan 14 10:36:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011221S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 13 03:57:57 2021
Response via : Initial Calibration

