

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012021\
 Data File : VD068105.D
 Acq On : 20 Jan 2021 11:34
 Operator : VA/SY
 Sample : VD0120SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0120SBL01

Quant Time: Jan 20 14:31:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	492944	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	849105	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	760467	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.578	152	345692	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	229899	48.75	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.50%
35) Dibromofluoromethane	7.920	113	226675	48.29	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.58%
50) Toluene-d8	10.343	98	857546	45.67	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.34%
62) 4-Bromofluorobenzene	12.631	95	316865	46.86	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.72%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012021\
Data File : VD068105.D
Acq On : 20 Jan 2021 11:34
Operator : VA/SY
Sample : VD0120SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0120SBL01

Quant Time: Jan 20 14:31:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

