

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012021\
 Data File : VD068108.D
 Acq On : 20 Jan 2021 13:03
 Operator : VA/SY
 Sample : M1160-03
 Misc : 4.90G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BASIN-1

Quant Time: Jan 20 14:34:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	152746	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	265849	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	274360	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	135479	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	92020	62.98	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 125.96%	
35) Dibromofluoromethane	7.920	113	76640	52.14	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.28%	
50) Toluene-d8	10.344	98	339212	57.70	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 115.40%	
62) 4-Bromofluorobenzene	12.632	95	138085	65.22	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 130.44%	
Target Compounds						
						Qvalue
16) Acetone	4.162	43	104037	398.76	ug/l	96
25) 2-Butanone	7.215	43	61621	156.86	ug/l	89
59) 2-Hexanone	10.985	43	5368	8.43	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012021\
Data File : VD068108.D
Acq On : 20 Jan 2021 13:03
Operator : VA/SY
Sample : M1160-03
Misc : 4.90G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BASIN-1

Quant Time: Jan 20 14:34:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

