

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012021\
 Data File : VD068112.D
 Acq On : 20 Jan 2021 14:52
 Operator : VA/SY
 Sample : M1160-05RE
 Misc : 3.85G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BASIN-2RE

Quant Time: Jan 20 15:12:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	119930	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	223141	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	243828	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	129289	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	87087	75.91	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	151.82%
35) Dibromofluoromethane	7.920	113	67545	54.75	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.50%
50) Toluene-d8	10.344	98	286138	57.99	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.98%
62) 4-Bromofluorobenzene	12.632	95	129419	72.83	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	145.66%#
Target Compounds						
					Qvalue	
16) Acetone	4.167	43	5787	28.25	ug/l	97
18) Methyl Acetate	4.720	43	7924	14.90	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012021\
Data File : VD068112.D
Acq On : 20 Jan 2021 14:52
Operator : VA/SY
Sample : M1160-05RE
Misc : 3.85G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BASIN-2RE

Quant Time: Jan 20 15:12:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

