

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012021\
 Data File : VD068116.D
 Acq On : 20 Jan 2021 16:40
 Operator : VA/SY
 Sample : M1082-03
 Misc : 6.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BACKFILL-ASHCROFT-02

Quant Time: Jan 20 23:46:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	423610	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	742560	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	655235	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	290222	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	201724	49.78	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.56%
35) Dibromofluoromethane	7.920	113	191130	46.56	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.12%
50) Toluene-d8	10.344	98	721924	43.96	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	87.92%
62) 4-Bromofluorobenzene	12.632	95	263192	44.51	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.02%
Target Compounds						
16) Acetone	4.161	43	23833	32.94	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012021\
Data File : VD068116.D
Acq On : 20 Jan 2021 16:40
Operator : VA/SY
Sample : M1082-03
Misc : 6.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BACKFILL-ASHCROFT-02

Quant Time: Jan 20 23:46:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

