

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012021\
 Data File : VD068117.D
 Acq On : 20 Jan 2021 17:07
 Operator : VA/SY
 Sample : M1100-02
 Misc : 6.70G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TOPSOIL-RBF-ASHCROFT

Quant Time: Jan 20 23:46:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	404986	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	703648	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	591633	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	221057	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	203613	52.56	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.12%
35) Dibromofluoromethane	7.920	113	192119	49.39	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.78%
50) Toluene-d8	10.344	98	681614	43.80	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	87.60%
62) 4-Bromofluorobenzene	12.632	95	227660	40.63	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.26%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012021\
Data File : VD068117.D
Acq On : 20 Jan 2021 17:07
Operator : VA/SY
Sample : M1100-02
Misc : 6.70G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TOPSOIL-RBF-ASHCROFT

Quant Time: Jan 20 23:46:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

