

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012021\
 Data File : VD068118.D
 Acq On : 20 Jan 2021 17:34
 Operator : VA/SY
 Sample : M1100-03
 Misc : 6.39G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TOPSOIL-RBF-DARK

Quant Time: Jan 20 23:47:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	396994	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	682541	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	561582	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.579	152	206669	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	194456	51.20	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.40%	
35) Dibromofluoromethane	7.920	113	191281	50.69	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.38%	
50) Toluene-d8	10.344	98	701129	46.45	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 92.90%	
62) 4-Bromofluorobenzene	12.632	95	220552	40.58	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 81.16%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012021\
Data File : VD068118.D
Acq On : 20 Jan 2021 17:34
Operator : VA/SY
Sample : M1100-03
Misc : 6.39G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TOPSOIL-RBF-DARK

Quant Time: Jan 20 23:47:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

