

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012121\
 Data File : VD068131.D
 Acq On : 21 Jan 2021 14:31
 Operator : VA/SY
 Sample : M1152-02
 Misc : 7.71G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-EF-01-017-ABCD

Quant Time: Jan 22 00:09:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	575781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	976232	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.650	117	968113	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	457494	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	261224	47.43	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.86%
35) Dibromofluoromethane	7.920	113	274412	50.84	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.68%
50) Toluene-d8	10.344	98	1228646	56.91	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	113.82%
62) 4-Bromofluorobenzene	12.632	95	465498	59.88	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	119.76%

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012121\
Data File : VD068131.D
Acq On : 21 Jan 2021 14:31
Operator : VA/SY
Sample : M1152-02
Misc : 7.71G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-EF-01-017-ABCD

Quant Time: Jan 22 00:09:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

